

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
 Data File : BF128040.D
 Acq On : 03 May 2022 12:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 04:25:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	69	0.00
2	1,4-Dioxane	40.000	38.704	3.2	65	0.00
3	Pyridine	40.000	40.556	-1.4	67	0.00
4	n-Nitrosodimethylamine	40.000	42.411	-6.0	71	0.00
5 S	2-Fluorophenol	80.000	80.807	-1.0	70	0.00
6	Aniline	40.000	41.546	-3.9	69	0.00
7 S	Phenol-d6	80.000	82.971	-3.7	71	0.00
8	2-Chlorophenol	40.000	41.092	-2.7	69	0.00
9	Benzaldehyde	40.000	34.924	12.7	68	0.00
10 C	Phenol	40.000	42.718	-6.8	72	0.00
11	bis(2-Chloroethyl)ether	40.000	40.717	-1.8	69	0.00
12	1,3-Dichlorobenzene	40.000	39.930	0.2	69	0.00
13 C	1,4-Dichlorobenzene	40.000	40.039	-0.1	69	0.00
14	1,2-Dichlorobenzene	40.000	40.870	-2.2	71	0.00
15	Benzyl Alcohol	40.000	51.304	-28.3#	82	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.837	-2.1	68	0.00
17	2-Methylphenol	40.000	40.677	-1.7	67	0.00
18	Hexachloroethane	40.000	42.012	-5.0	72	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.613	-6.5	73	0.00
20	3+4-Methylphenols	40.000	41.513	-3.8	70	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	67	0.00
22	Acetophenone	40.000	42.452	-6.1	72	0.00
23 S	Nitrobenzene-d5	80.000	89.192	-11.5	74	0.00
24	Nitrobenzene	40.000	43.397	-8.5	72	0.00
25	Isophorone	40.000	43.096	-7.7	71	0.00
26 C	2-Nitrophenol	40.000	45.310	-13.3	72	0.00
27	2,4-Dimethylphenol	40.000	41.743	-4.4	67	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.654	-6.6	71	0.00
29 C	2,4-Dichlorophenol	40.000	42.073	-5.2	70	0.00
30	1,2,4-Trichlorobenzene	40.000	41.784	-4.5	70	0.00
31	Naphthalene	40.000	41.273	-3.2	70	0.00
32	Benzoic acid	40.000	41.828	-4.6	63	0.00
33	4-Chloroaniline	40.000	41.577	-3.9	68	0.00
34 C	Hexachlorobutadiene	40.000	45.134	-12.8	76	0.00
35	Caprolactam	40.000	41.837	-4.6	67	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.439	-8.6	71	0.00
37	2-Methylnaphthalene	40.000	42.222	-5.6	71	0.00
38	1-Methylnaphthalene	40.000	42.293	-5.7	71	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	69	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.095	-7.7	74	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.490	-3.7	67	0.00
42 S	2,4,6-Tribromophenol	80.000	94.992	-18.7	81	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.552	-6.4	72	0.00
44	2,4,5-Trichlorophenol	40.000	41.236	-3.1	68	0.00
45 S	2-Fluorobiphenyl	80.000	90.852	-13.6	77	0.00
46	1,1'-Biphenyl	40.000	42.145	-5.4	73	0.00
47	2-Chloronaphthalene	40.000	41.546	-3.9	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
 Data File : BF128040.D
 Acq On : 03 May 2022 12:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 04:25:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.598	-14.0	75	0.00
49	Acenaphthylene	40.000	41.593	-4.0	71	0.00
50	Dimethylphthalate	40.000	42.294	-5.7	72	0.00
51	2,6-Dinitrotoluene	40.000	42.746	-6.9	70	0.00
52 C	Acenaphthene	40.000	42.159	-5.4	71	0.00
53	3-Nitroaniline	40.000	41.041	-2.6	67	0.00
54 P	2,4-Dinitrophenol	40.000	45.635	-14.1	71	0.00
55	Dibenzofuran	40.000	41.493	-3.7	71	0.00
56 P	4-Nitrophenol	40.000	38.232	4.4	61	0.00
57	2,4-Dinitrotoluene	40.000	44.311	-10.8	72	0.00
58	Fluorene	40.000	42.573	-6.4	74	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.559	-6.4	71	0.00
60	Diethylphthalate	40.000	43.334	-8.3	73	0.00
61	4-Chlorophenyl-phenylether	40.000	43.338	-8.3	75	0.00
62	4-Nitroaniline	40.000	41.599	-4.0	68	0.00
63	Azobenzene	40.000	43.227	-8.1	73	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	69	0.00
65	4,6-Dinitro-2-methylphenol	40.000	48.318	-20.8	73	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.178	-2.9	72	0.00
67	4-Bromophenyl-phenylether	40.000	42.771	-6.9	74	0.00
68	Hexachlorobenzene	40.000	43.729	-9.3	76	0.00
69	Atrazine	40.000	41.768	-4.4	73	0.00
70 C	Pentachlorophenol	40.000	38.108	4.7	61	0.00
71	Phenanthrene	40.000	40.809	-2.0	71	0.00
72	Anthracene	40.000	41.201	-3.0	72	0.00
73	Carbazole	40.000	39.795	0.5	69	0.00
74	Di-n-butylphthalate	40.000	42.434	-6.1	73	0.00
75 C	Fluoranthene	40.000	40.585	-1.5	71	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	67	0.00
77	Benzydine	40.000	53.448	-33.6#	83	0.00
78	Pyrene	40.000	41.583	-4.0	70	0.00
79 S	Terphenyl-d14	80.000	88.786	-11.0	77	0.00
80	Butylbenzylphthalate	40.000	44.639	-11.6	72	0.00
81	Benzo(a)anthracene	40.000	41.633	-4.1	69	0.00
82	3,3'-Dichlorobenzidine	40.000	38.012	5.0	64	0.00
83	Chrysene	40.000	40.249	-0.6	68	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.892	-9.7	72	0.00
85 c	Di-n-octyl phthalate	40.000	42.723	-6.8	69	0.00
86 I	Perylene-d12	20.000	20.000	0.0	64	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.854	-4.6	63	0.00
88	Benzo(b)fluoranthene	40.000	41.986	-5.0	66	0.00
89	Benzo(k)fluoranthene	40.000	40.814	-2.0	62	0.00
90 C	Benzo(a)pyrene	40.000	41.730	-4.3	64	0.00
91	Dibenzo(a,h)anthracene	40.000	42.227	-5.6	63	0.00
92	Benzo(g,h,i)perylene	40.000	41.781	-4.5	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
Data File : BF128040.D
Acq On : 03 May 2022 12:17
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 04 04:25:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 04:23:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0