

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050416\
 Data File : BF086688.D
 Acq On : 4 May 2016 20:39
 Operator : UM/SJ
 Sample : H2715-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 26WARD-3-CS-3(0-12FTBGS)

Manual Integrations
 APPROVED

sohil
 5/5/2016 3:19:01 PM

Quant Time: May 05 04:51:13 2016
 Quant Title :
 QLast Update : Wed May 04 15:12:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	157710	40.00	ng	0.00
21) Naphthalene-d8	8.02	136	559316	40.00	ng	-0.01
38) Acenaphthene-d10	9.78	164	236151	40.00	ng	0.00
63) Phenanthrene-d10	11.25	188	339591m	40.00	ng	0.00
75) Chrysene-d12	13.88	240	299816	40.00	ng	0.00
86) Perylene-d12	15.28	264	272778	40.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	1212288	127.25	ng	0.01
7) Phenol-d6	6.38	99	1520804	118.51	ng	0.01
23) Nitrobenzene-d5	7.31	82	1170377	107.53	ng	0.00
41) 2,4,6-Tribromophenol	10.57	330	305261	132.14	ng	0.00
44) 2-Fluorobiphenyl	9.10	172	1333691	92.58	ng	0.00
78) Terphenyl-d14	12.84	244	806180	54.34	ng	0.00
Target Compounds						Ovalue
49) Dimethylphthalate	9.49	163	203029	11.53	ng	100
70) Phenanthrene	11.28	178	359610	19.71	ng	100
71) Anthracene	11.33	178	89939	4.71	ng	99
72) Carbazole	11.49	167	36529	2.20	ng	96
74) Fluoranthene	12.46	202	481094	26.32	ng	95
77) Pyrene	12.69	202	422343	17.67	ng	99
80) Benzo(a)anthracene	13.88	228	206751	12.19	ng	99
82) Chrysene	13.92	228	179415m	9.92	ng	
85) Indeno(1,2,3-cd)pyrene	16.59	276	72597	4.68	ng	# 91
87) Benzo(b)fluoranthene	14.89	252	190319m	11.61	ng	
88) Benzo(k)fluoranthene	14.90	252	95641m	6.19	ng	
89) Benzo(a)pyrene	15.22	252	143998	9.54	ng	# 92
91) Benzo(a,h,i)perylene	16.99	276	63490	4.17	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050416\
Data File : BF086688.D
Acq On : 4 May 2016 20:39
Operator : UM/SJ
Sample : H2715-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
26WARD-3-CS-3(0-12FTBGS)

Manual Integrations
APPROVED

sohil
5/5/2016 3:19:01 PM

Quant Time: May 05 04:51:13 2016
Quant Title :
QLast Update : Wed May 04 15:12:53 2016
Response via : Initial Calibration

