

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050421\
 Data File : BF123820.D
 Acq On : 4 May 2021 21:28
 Operator : JU/CG
 Sample : M2254-05
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-2

Quant Time: May 05 05:14:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 04 12:58:27 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	254228	20.00	ng	-0.01
21) Naphthalene-d8	8.086	136	954410	20.00	ng	-0.01
39) Acenaphthene-d10	9.851	164	359055	20.00	ng	0.00
64) Phenanthrene-d10	11.357	188	489581	20.00	ng	# 0.02
76) Chrysene-d12	13.992	240	442406	20.00	ng	# 0.00
86) Perylene-d12	15.439	264	543162	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	1342297	83.65	ng	0.00
7) Phenol-d6	6.439	99	1780314	80.54	ng	0.00
23) Nitrobenzene-d5	7.363	82	1252918	59.26	ng	-0.01
42) 2,4,6-Tribromophenol	10.651	330	355787	79.60	ng	0.00
45) 2-Fluorobiphenyl	9.163	172	1802076	73.09	ng	-0.01
79) Terphenyl-d14	12.945	244	1198755	52.18	ng	0.02
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.363	70	169955	12.45	ng	# 74
25) Isophorone	7.363	82	1252938	31.61	ng	# 69
32) Benzoic acid	7.881	122	519	5.44	ng	# 28
50) Dimethylphthalate	9.557	163	63800	2.52	ng	# 1
51) 2,6-Dinitrotoluene	9.651	165	28902	4.94	ng	# 19
53) 3-Nitroaniline	9.804	138	66310	9.50	ng	# 10
57) 2,4-Dinitrotoluene	10.057	165	18610	2.33	ng	# 52
58) Fluorene	10.404	166	60256	2.38	ng	# 47
61) 4-Chlorophenyl-phenyle...	10.422	204	29040	2.46	ng	# 1
62) 4-Nitroaniline	10.198	138	38523	5.57	ng	85
66) n-Nitrosodiphenylamine	10.522	169	118525	7.40	ng	# 48
69) Atrazine	11.033	200	11264	2.33	ng	# 1
71) Phenanthrene	11.380	178	336080	13.09	ng	# 72
72) Anthracene	11.433	178	84364	3.30	ng	# 41
73) Carbazole	11.580	167	53320	2.07	ng	# 1
74) Di-n-butylphthalate	11.916	149	122163	4.22	ng	# 37
75) Fluoranthene	12.586	202	562767	20.09	ng	94
77) Benzidine	12.651	184	23484	2.05	ng	# 1
78) Pyrene	12.810	202	503210	14.65	ng	96
81) Benzo(a)anthracene	13.980	228	286644	10.69	ng	96
83) Chrysene	13.980	228	286644	10.63	ng	96
87) Indeno(1,2,3-cd)pyrene	16.874	276	198700	6.28	ng	98
88) Benzo(b)fluoranthene	15.021	252	558686	15.26	ng	# 95
89) Benzo(k)fluoranthene	15.021	252	558686	15.99	ng	# 95
90) Benzo(a)pyrene	15.374	252	306016	9.78	ng	# 94
92) Benzo(g,h,i)perylene	17.309	276	180891	6.76	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050421\
Data File : BF123820.D
Acq On : 4 May 2021 21:28
Operator : JU/CG
Sample : M2254-05
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-2

Quant Time: May 05 05:14:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 04 12:58:27 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.798 min Scan# 801
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampleId :
TP-2


Tgt Ion:152 Resp: 254228
Ion Ratio Lower Upper
152 100
150 160.1 127.0 190.6
115 64.7 54.3 81.5



#5
2-Fluorophenol
Concen: 83.65 ng
RT: 5.422 min Scan# 567
Delta R.T. -0.000 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:112 Resp: 1342297
Ion Ratio Lower Upper
112 100
64 75.2 61.4 92.0
63 40.5 33.0 49.4









#7
Phenol-d6
Concen: 80.54 ng
RT: 6.439 min Scan# 740
Delta R.T. -0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Instrument :
BNA_F
ClientSampleId :
TP-2

Tgt Ion: 99 Resp: 1780314




Ion	Ratio	Lower	Upper
99	100		
42	24.2	19.8	29.8
71	39.7	31.1	46.7




#19
n-Nitroso-di-n-propylamine
Concen: 12.45 ng
RT: 7.363 min Scan# 897
Delta R.T. 0.135 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion: 70 Resp: 169955

Ion	Ratio	Lower	Upper
70	100		
42	54.7	60.1	90.1#
101	0.0	7.4	11.2#
130	1.4	14.2	21.2#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.086 min Scan# 1020
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Instrument :
BNA_F
ClientSampled :
TP-2

Tgt Ion:136 Resp: 954410




Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	11.2	9.8	14.6
68	5.7	4.1	6.1




#23
Nitrobenzene-d5
Concen: 59.26 ng
RT: 7.363 min Scan# 897
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion: 82 Resp: 1252918

Ion	Ratio	Lower	Upper
82	100		
128	37.6	29.4	44.2
54	65.7	54.1	81.1









#25
Isophorone
Concen: 31.61 ng
RT: 7.363 min Scan# 897
Delta R.T. -0.271 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion: 82 Resp: 1252938
Ion Ratio Lower Upper
82 100
95 0.0 5.8 8.6#
138 0.0 11.8 17.8#



#32
Benzoic acid
Concen: 5.44 ng
RT: 7.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:122 Resp: 519
Ion Ratio Lower Upper
122 100
105 0.0 107.8 147.8#
77 78.2 81.6 121.6#









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.851 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:164 Resp: 359055
Ion Ratio Lower Upper
164 100
162 99.6 80.7 121.1
160 45.1 37.3 55.9



#42
2,4,6-Tribromophenol
Concen: 79.60 ng
RT: 10.651 min Scan# 1456
Delta R.T. 0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:330 Resp: 355787
Ion Ratio Lower Upper
330 100
332 96.6 77.1 115.7
141 26.0 27.9 41.9#









#45
2-Fluorobiphenyl
Concen: 73.09 ng
RT: 9.163 min Scan# 1203
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:172 Resp: 1802076
Ion Ratio Lower Upper
172 100
171 34.0 28.1 42.1
170 23.1 19.2 28.8



#50
Dimethylphthalate
Concen: 2.52 ng
RT: 9.557 min Scan# 1270
Delta R.T. -0.024 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:163 Resp: 63800
Ion Ratio Lower Upper
163 100
194 128.6 3.1 4.7#
164 19.7 8.2 12.2#









#51
2,6-Dinitrotoluene
Concen: 4.94 ng
RT: 9.651 min Scan# 1286
Delta R.T. 0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampleId :
TP-2


Tgt Ion:165 Resp: 28902
Ion Ratio Lower Upper
165 100
63 9.6 67.7 101.5#
89 6.1 58.8 88.2#



#53
3-Nitroaniline
Concen: 9.50 ng
RT: 9.804 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:138 Resp: 66310
Ion Ratio Lower Upper
138 100
108 50.8 10.2 15.4#
92 25.1 103.3 154.9#









#57
2,4-Dinitrotoluene
Concen: 2.33 ng
RT: 10.057 min Scan# 1355
Delta R.T. 0.018 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:165 Resp: 18610
Ion Ratio Lower Upper
165 100
63 41.0 43.5 65.3#
89 22.7 67.7 101.5#
182 14.7 2.1 3.1#



#58
Fluorene
Concen: 2.38 ng
RT: 10.404 min Scan# 1414
Delta R.T. -0.000 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:166 Resp: 60256
Ion Ratio Lower Upper
166 100
165 140.2 73.5 110.3#
167 40.5 10.7 16.1#









#61
4-Chlorophenyl-phenylether
Concen: 2.46 ng
RT: 10.422 min Scan# 1417
Delta R.T. 0.029 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:204 Resp: 29040
Ion Ratio Lower Upper
204 100
206 76.7 25.7 38.5#
141 233.4 55.5 83.3#



#62
4-Nitroaniline
Concen: 5.57 ng
RT: 10.198 min Scan# 1379
Delta R.T. -0.224 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:138 Resp: 38523
Ion Ratio Lower Upper
138 100
92 52.0 35.8 75.8
108 103.1 63.6 103.6









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.357 min Scan# 1576
Delta R.T. 0.018 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:188 Resp: 489581
Ion Ratio Lower Upper
188 100
94 23.5 6.9 10.3#
80 23.2 8.2 12.2#



#66
n-Nitrosodiphenylamine
Concen: 7.40 ng
RT: 10.522 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:169 Resp: 118525
Ion Ratio Lower Upper
169 100
168 17.5 53.3 79.9#
167 16.9 29.9 44.9#





#69

Atrazine

Concen: 2.33 ng

RT: 11.033 min Scan# 1521

Delta R.T. -0.006 min

Lab File: BF123820.D

Acq: 4 May 2021 21:28





Tgt Ion:200 Resp: 11264

Ion Ratio Lower Upper

200 100

173 306.6 6.9 46.9#

215 325.0 26.5 66.5#



#71

Phenanthrene

Concen: 13.09 ng

RT: 11.380 min Scan# 1580

Delta R.T. 0.018 min

Lab File: BF123820.D

Acq: 4 May 2021 21:28




Tgt Ion:178 Resp: 336080


Ion Ratio Lower Upper

178 100

176 21.4 15.6 23.4

179 40.1 12.5 18.7#









#72
 Anthracene
 Concen: 3.30 ng
 RT: 11.433 min Scan# 1589
 Delta R.T. 0.018 min
 Lab File: BF123820.D
 Acq: 4 May 2021 21:28




Instrument :
 BNA_F
 ClientSampled :
 TP-2


Tgt Ion:178 Resp: 84364
 Ion Ratio Lower Upper
 178 100
 176 30.5 15.4 23.0#
 179 58.0 12.6 18.8#



#73
 Carbazole
 Concen: 2.07 ng
 RT: 11.580 min Scan# 1614
 Delta R.T. 0.012 min
 Lab File: BF123820.D
 Acq: 4 May 2021 21:28

Tgt Ion:167 Resp: 53320
 Ion Ratio Lower Upper
 167 100
 166 64.2 17.0 25.6#
 139 90.0 10.4 15.6#









#74
Di-n-butylphthalate
Concen: 4.22 ng
RT: 11.916 min Scan# 1671
Delta R.T. 0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampleId :
TP-2


Tgt Ion:149 Resp: 122163
Ion Ratio Lower Upper
149 100
150 42.7 7.8 11.8#
104 12.0 4.6 6.8#



#75
Fluoranthene
Concen: 20.09 ng
RT: 12.586 min Scan# 1785
Delta R.T. 0.029 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:202 Resp: 562767
Ion Ratio Lower Upper
202 100
101 11.9 0.0 30.9
203 21.3 0.0 37.5









#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.992 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:240 Resp: 442406
Ion Ratio Lower Upper
240 100
120 12.2 7.2 10.8#
236 25.3 19.0 28.4



#77
Benzidine
Concen: 2.05 ng
RT: 12.651 min Scan# 1796
Delta R.T. -0.024 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:184 Resp: 23484
Ion Ratio Lower Upper
184 100
185 118.5 12.1 18.1#
183 376.2 7.8 11.8#









#78
Pyrene
Concen: 14.65 ng
RT: 12.810 min Scan# 1823
Delta R.T. 0.029 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampleId :
TP-2


Tgt Ion:202 Resp: 503210
Ion Ratio Lower Upper
202 100
200 20.3 16.1 24.1
203 21.9 14.8 22.2



#79
Terphenyl-d14
Concen: 52.18 ng
RT: 12.945 min Scan# 1846
Delta R.T. 0.018 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:244 Resp: 1198755
Ion Ratio Lower Upper
244 100
212 7.0 5.7 8.5
122 13.8 9.7 14.5









#81
Benzo(a)anthracene
Concen: 10.69 ng
RT: 13.980 min Scan# 2022
Delta R.T. 0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:228 Resp: 286644
Ion Ratio Lower Upper
228 100
226 29.1 22.4 33.6
229 23.1 16.0 24.0



#83
Chrysene
Concen: 10.63 ng
RT: 13.980 min Scan# 2022
Delta R.T. -0.030 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:228 Resp: 286644
Ion Ratio Lower Upper
228 100
226 29.1 24.3 36.5
229 23.1 16.2 24.2









#86
Perylene-d12
Concen: 20.00 ng
RT: 15.439 min Scan# 2270
Delta R.T. -0.000 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampleId :
TP-2


Tgt Ion:264 Resp: 543162
Ion Ratio Lower Upper
264 100
260 24.3 18.7 28.1
265 22.1 17.8 26.8



#87
Indeno(1,2,3-cd)pyrene
Concen: 6.28 ng
RT: 16.874 min Scan# 2514
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:276 Resp: 198700
Ion Ratio Lower Upper
276 100
138 19.4 16.2 24.2
277 24.1 20.1 30.1









#88
Benzo(b)fluoranthene
Concen: 15.26 ng
RT: 15.021 min Scan# 2199
Delta R.T. 0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28




Instrument :
BNA_F
ClientSampled :
TP-2


Tgt Ion:252 Resp: 558686
Ion Ratio Lower Upper
252 100
253 22.8 18.1 27.1
125 15.0 6.2 9.4#



#89
Benzo(k)fluoranthene
Concen: 15.99 ng
RT: 15.021 min Scan# 2199
Delta R.T. -0.024 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:252 Resp: 558686
Ion Ratio Lower Upper
252 100
253 22.8 17.9 26.9
125 15.0 7.0 10.6#









#90
Benzo(a)pyrene
Concen: 9.78 ng
RT: 15.374 min Scan# 2259
Delta R.T. -0.006 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Instrument :
BNA_F
ClientSampleId :
TP-2

Tgt Ion:252 Resp: 306016
Ion Ratio Lower Upper
252 100
253 23.8 18.4 27.6
125 16.3 8.0 12.0#



#92
Benzo(g,h,i)perylene
Concen: 6.76 ng
RT: 17.309 min Scan# 2588
Delta R.T. -0.012 min
Lab File: BF123820.D
Acq: 4 May 2021 21:28

Tgt Ion:276 Resp: 180891
Ion Ratio Lower Upper
276 100
277 25.7 18.2 27.2
138 21.7 13.8 20.6#

