

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050421\
 Data File : BF123808.D
 Acq On : 4 May 2021 14:55
 Operator : JU/CG
 Sample : PB136158TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB136158TB

Quant Time: May 04 15:31:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 04 12:58:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	273201	20.00	ng	0.00
21) Naphthalene-d8	8.081	136	1038576	20.00	ng	0.00
39) Acenaphthene-d10	9.839	164	514628	20.00	ng	0.00
64) Phenanthrene-d10	11.327	188	1016865	20.00	ng	0.00
76) Chrysene-d12	13.968	240	883507	20.00	ng	0.00
86) Perylene-d12	15.421	264	715978	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	2105304	122.08	ng	0.01
7) Phenol-d6	6.445	99	2740600	115.37	ng	0.00
23) Nitrobenzene-d5	7.363	82	1952669	84.87	ng	0.00
42) 2,4,6-Tribromophenol	10.633	330	820007	128.00	ng	0.00
45) 2-Fluorobiphenyl	9.163	172	2998179	84.84	ng	0.00
79) Terphenyl-d14	12.921	244	3588555	78.21	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050421\
Data File : BF123808.D
Acq On : 4 May 2021 14:55
Operator : JU/CG
Sample : PB136158TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB136158TB

Quant Time: May 04 15:31:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 04 12:58:27 2021
Response via : Initial Calibration

