

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050516\
 Data File : BF086708.D
 Acq On : 5 May 2016 14:31
 Operator : UM/SJ
 Sample : PB90099TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB90099TB

Quant Time: May 06 01:25:33 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:12:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	174265	40.00	ng	-0.01
21) Naphthalene-d8	8.01	136	742370	40.00	ng	-0.02
38) Acenaphthene-d10	9.76	164	343601	40.00	ng	-0.02
63) Phenanthrene-d10	11.24	188	695987	40.00	ng	-0.01
75) Chrysene-d12	13.87	240	566763	40.00	ng	-0.01
86) Perylene-d12	15.25	264	459263	40.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	1235576	117.38	ng	0.00
7) Phenol-d6	6.37	99	1827334	128.87	ng	0.00
23) Nitrobenzene-d5	7.29	82	1054158	72.97	ng	-0.02
41) 2,4,6-Tribromophenol	10.56	330	426209	126.80	ng	-0.01
44) 2-Fluorobiphenyl	9.09	172	1902429	90.76	ng	-0.01
78) Terphenyl-d14	12.83	244	1638403	58.42	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050516\
Data File : BF086708.D
Acq On : 5 May 2016 14:31
Operator : UM/SJ
Sample : PB90099TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB90099TB

Quant Time: May 06 01:25:33 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 15:12:53 2016
Response via : Initial Calibration

