

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050516\
 Data File : BF086710.D
 Acq On : 5 May 2016 15:28
 Operator : UM/SJ
 Sample : H2804-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FILTER-CAKE

Quant Time: May 06 01:31:05 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:12:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	195579	40.00	ng	-0.01
21) Naphthalene-d8	8.01	136	799782	40.00	ng	-0.02
38) Acenaphthene-d10	9.76	164	378394	40.00	ng	-0.02
63) Phenanthrene-d10	11.24	188	768100	40.00	ng	-0.01
75) Chrysene-d12	13.87	240	601802	40.00	ng	-0.01
86) Perylene-d12	15.25	264	487332	40.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	1116793	94.53	ng	0.01
7) Phenol-d6	6.37	99	1566001	98.40	ng	0.00
23) Nitrobenzene-d5	7.30	82	1605161	103.14	ng	-0.01
41) 2,4,6-Tribromophenol	10.56	330	492775	133.13	ng	-0.01
44) 2-Fluorobiphenyl	9.09	172	2307384	99.96	ng	-0.01
78) Terphenyl-d14	12.83	244	2117979	71.13	ng	-0.01
Target Compounds						
31) Naphthalene	8.03	128	196018	4.80	ng	99
32) Benzoic acid	7.76	122	16189	11.13	ng	# 80
69) Pentachlorophenol	11.05	266	3406	6.52	ng	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050516\
Data File : BF086710.D
Acq On : 5 May 2016 15:28
Operator : UM/SJ
Sample : H2804-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FILTER-CAKE

Quant Time: May 06 01:31:05 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 15:12:53 2016
Response via : Initial Calibration

