

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050517\
 Data File : BF094922.D
 Acq On : 5 May 2017 18:00
 Operator : SJ/MA
 Sample : I2930-01DL 10X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RO-1-COMPDL

Manual Integrations
 APPROVED

mohammad
 5/8/2017 2:47:50 PM

Quant Time: May 05 18:40:59 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 17:24:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	90587	20.00	ng	-0.01
21) Naphthalene-d8	9.74	136	358455	20.00	ng	-0.01
38) Acenaphthene-d10	12.57	164	172131	20.00	ng	-0.01
63) Phenanthrene-d10	14.98	188	287191	20.00	ng	0.00
75) Chrysene-d12	18.64	240	281653	20.00	ng	-0.02
86) Perylene-d12	20.30	264	192366	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.69	112	54099m	9.96	ng	0.01
7) Phenol-d6	7.28	99	69345	10.64	ng	0.00
23) Nitrobenzene-d5	8.62	82	41320	7.27	ng	-0.02
41) 2,4,6-Tribromophenol	13.88	330	13613	9.66	ng	0.00
44) 2-Fluorobiphenyl	11.51	172	91059	7.61	ng	-0.02
78) Terphenyl-d14	17.29	244	83462	6.53	ng	-0.02
Target Compounds						
31) Naphthalene	9.77	128	65636	3.64	ng	# 92
37) 2-Methylnaphthalene	10.91	142	710815	60.12	ng	98
57) Fluorene	13.47	166	80200	6.09	ng	# 98
70) Phenanthrene	15.01	178	281644	17.07	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050517\
Data File : BF094922.D
Acq On : 5 May 2017 18:00
Operator : SJ/MA
Sample : I2930-01DL 10X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RO-1-COMPDL

Manual Integrations
APPROVED

mohammad
5/8/2017 2:47:50 PM

Quant Time: May 05 18:40:59 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 17:24:51 2017
Response via : Initial Calibration

