



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050522\
 Data File : BF128090.D
 Acq On : 05 May 2022 13:07
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 05 17:56:23 2022






AutoFind: Scans 1596, 1597, 1598; Background Corrected with Scan 1591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	44547	PASS
68	69	0.00	2	1.8	795	PASS
69	198	0.00	100	38.2	43899	PASS
70	69	0.00	2	0.4	174	PASS
127	198	10	80	45.6	52403	PASS
197	198	0.00	2	0.2	223	PASS
198	198	100	100	100.0	115027	PASS
199	198	5	9	7.1	8212	PASS
275	198	10	60	27.8	31965	PASS
365	198	1	100	3.9	4495	PASS
441	198	0.01	100	15.8	18156	PASS
442	442	50	100	100.0	115808	PASS
443	442	15	24	18.9	21921	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050522\
Data File : BF128090.D
Acq On : 05 May 2022 13:07
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 05 18:14:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 17:56:23 2022
Response via : Initial Calibration



TIC: BF128090.D\data.ms

(70) Pentachlorophenol (C)

11.239min (-0.006) 24206.75 ng




response 88154

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.30	63.59
264.00	66.40	61.20
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050522\
Data File : BF128090.D
Acq On : 05 May 2022 13:07
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 05 18:14:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 17:56:23 2022
Response via : Initial Calibration



TIC: BF128090.D\data.ms

(77) Benzidine

12.780min (+ 0.000) 0.00 ng

response 516502

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.10	14.01
183.00	11.20	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/5/2022	BNA_F	<u>BF128090.D</u>
Compound Name	Response	Retention Time
DDT	270822	13.59
DDD	3621	13.29
DDE	370	12.96
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3991	274813	1.45

Instrument :
BNA_F
ClientSampleId :
DFTPP