

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050615\
 Data File : BF078959.D
 Acq On : 6 May 2015 14:39
 Operator : TP/IZ
 Sample : G2116-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-47D

Manual Integrations
 APPROVED

apatel
 5/7/2015 1:57:06 PM

Quant Time: May 07 01:42:17 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 07 00:45:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.35	152	77039	20.00	ng	0.00
21) Naphthalene-d8	8.92	136	327928	20.00	ng	0.00
38) Acenaphthene-d10	11.10	164	154462	20.00	ng	0.00
63) Phenanthrene-d10	12.95	188	304011	20.00	ng	0.00
75) Chrysene-d12	16.27	240	316316	20.00	ng	-0.03
86) Perylene-d12	18.14	264	344281	20.00	ng	-0.09
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	426048	95.20	ng	0.01
7) Phenol-d6	6.90	99	574150	97.05	ng	0.01
23) Nitrobenzene-d5	8.03	82	302061	52.94	ng	0.00
41) 2,4,6-Tribromophenol	12.08	330	145371	87.00	ng	0.00
44) 2-Fluorobiphenyl	10.27	172	547359	49.37	ng	0.00
78) Terphenyl-d14	14.95	244	602930	45.63	ng	0.00
Target Compounds						
10) Phenol	6.91	94	17271	2.52	ng	Qvalue 98
49) Dimethylphthalate	10.78	163	41951	3.70	ng	99
74) Fluoranthene	14.46	202	75144	4.24	ng	94
77) Pyrene	14.73	202	70685	3.88	ng	# 92
80) Benzo(a)anthracene	16.26	228	44524	2.57	ng	97
82) Chrysene	16.31	228	36248	2.18	ng	97
87) Benzo(b)fluoranthene	17.66	252	57660	3.06	ng	99
89) Benzo(a)pyrene	18.07	252	40433	2.33	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050615\
Data File : BF078959.D
Acq On : 6 May 2015 14:39
Operator : TP/IZ
Sample : G2116-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-47D

Manual Integrations
APPROVED

apatel
5/7/2015 1:57:06 PM

Quant Time: May 07 01:42:17 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 07 00:45:56 2015
Response via : Initial Calibration

