

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050615\
 Data File : BF078962.D
 Acq On : 6 May 2015 16:05
 Operator : TP/IZ
 Sample : G2116-03 2X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-59BS

Manual Integrations
 APPROVED

apatel
 5/7/2015 1:57:28 PM

Quant Time: May 07 01:52:32 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 07 00:45:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.35	152	72902	20.00	ng	0.00
21) Naphthalene-d8	8.92	136	305614	20.00	ng	0.00
38) Acenaphthene-d10	11.10	164	142444	20.00	ng	0.00
63) Phenanthrene-d10	12.95	188	291221	20.00	ng	0.00
75) Chrysene-d12	16.30	240	319603	20.00	ng	-0.01
86) Perylene-d12	18.19	264	342208	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	209683	49.51	ng	0.01
7) Phenol-d6	6.89	99	271176	48.44	ng	0.00
23) Nitrobenzene-d5	8.03	82	156006	29.34	ng	0.00
41) 2,4,6-Tribromophenol	12.08	330	77508	50.30	ng	0.00
44) 2-Fluorobiphenyl	10.27	172	307174	30.04	ng	0.00
78) Terphenyl-d14	14.95	244	357700	26.80	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.78	163	26780	2.56	ng	99
70) Phenanthrene	12.97	178	34601	2.12	ng	99
74) Fluoranthene	14.46	202	57353	3.38	ng	94
77) Pyrene	14.74	202	53258	2.89	ng	98
87) Benzo(b)fluoranthene	17.69	252	39438m	2.11	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050615\
Data File : BF078962.D
Acq On : 6 May 2015 16:05
Operator : TP/IZ
Sample : G2116-03 2X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-59BS

Manual Integrations
APPROVED

apatel
5/7/2015 1:57:28 PM

Quant Time: May 07 01:52:32 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 07 00:45:56 2015
Response via : Initial Calibration

