

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050622\
 Data File : BF128116.D
 Acq On : 06 May 2022 10:08
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 16:06:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 17:56:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	68	0.00
2	1,4-Dioxane	40.000	39.668	0.8	65	-0.01
3	Pyridine	40.000	41.697	-4.2	67	-0.01
4	n-Nitrosodimethylamine	40.000	43.521	-8.8	71	-0.01
5 S	2-Fluorophenol	80.000	81.557	-1.9	69	0.00
6	Aniline	40.000	42.726	-6.8	69	0.00
7 S	Phenol-d6	80.000	84.947	-6.2	71	0.00
8	2-Chlorophenol	40.000	41.750	-4.4	69	0.00
9	Benzaldehyde	40.000	35.410	11.5	67	0.00
10 C	Phenol	40.000	42.601	-6.5	71	0.00
11	bis(2-Chloroethyl)ether	40.000	40.641	-1.6	68	0.00
12	1,3-Dichlorobenzene	40.000	40.768	-1.9	69	0.00
13 C	1,4-Dichlorobenzene	40.000	41.044	-2.6	69	0.00
14	1,2-Dichlorobenzene	40.000	40.859	-2.1	70	0.00
15	Benzyl Alcohol	40.000	51.252	-28.1#	80	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.590	1.0	64	0.00
17	2-Methylphenol	40.000	41.303	-3.3	66	0.00
18	Hexachloroethane	40.000	42.311	-5.8	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.016	-7.5	72	0.00
20	3+4-Methylphenols	40.000	41.867	-4.7	69	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	67	0.00
22	Acetophenone	40.000	42.673	-6.7	72	0.00
23 S	Nitrobenzene-d5	80.000	88.808	-11.0	73	0.00
24	Nitrobenzene	40.000	43.245	-8.1	71	0.00
25	Isophorone	40.000	42.431	-6.1	69	0.00
26 C	2-Nitrophenol	40.000	45.930	-14.8	72	0.00
27	2,4-Dimethylphenol	40.000	41.658	-4.1	66	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.886	-4.7	69	0.00
29 C	2,4-Dichlorophenol	40.000	42.212	-5.5	69	0.00
30	1,2,4-Trichlorobenzene	40.000	41.783	-4.5	70	0.00
31	Naphthalene	40.000	41.344	-3.4	69	0.00
32	Benzoic acid	40.000	41.297	-3.2	62	0.00
33	4-Chloroaniline	40.000	42.466	-6.2	69	0.00
34 C	Hexachlorobutadiene	40.000	44.016	-10.0	73	0.00
35	Caprolactam	40.000	42.465	-6.2	67	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.776	-9.4	71	0.00
37	2-Methylnaphthalene	40.000	41.732	-4.3	70	0.00
38	1-Methylnaphthalene	40.000	41.958	-4.9	70	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.316	-5.8	71	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.359	16.6	53	0.00
42 S	2,4,6-Tribromophenol	80.000	93.184	-16.5	79	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.078	-7.7	71	0.00
44	2,4,5-Trichlorophenol	40.000	40.995	-2.5	67	0.00
45 S	2-Fluorobiphenyl	80.000	90.249	-12.8	75	0.00
46	1,1'-Biphenyl	40.000	41.426	-3.6	70	0.00
47	2-Chloronaphthalene	40.000	40.986	-2.5	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050622\
 Data File : BF128116.D
 Acq On : 06 May 2022 10:08
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 16:06:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 17:56:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.967	-12.4	73	0.00
49	Acenaphthylene	40.000	41.394	-3.5	69	0.00
50	Dimethylphthalate	40.000	41.285	-3.2	69	0.00
51	2,6-Dinitrotoluene	40.000	42.490	-6.2	69	0.00
52 C	Acenaphthene	40.000	41.713	-4.3	69	0.00
53	3-Nitroaniline	40.000	41.690	-4.2	67	0.00
54 P	2,4-Dinitrophenol	40.000	41.471	-3.7	64	0.00
55	Dibenzofuran	40.000	40.925	-2.3	69	0.00
56 P	4-Nitrophenol	40.000	36.648	8.4	58	0.00
57	2,4-Dinitrotoluene	40.000	42.805	-7.0	68	0.00
58	Fluorene	40.000	42.269	-5.7	72	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.312	-5.8	69	0.00
60	Diethylphthalate	40.000	41.594	-4.0	69	0.00
61	4-Chlorophenyl-phenylether	40.000	42.531	-6.3	73	0.00
62	4-Nitroaniline	40.000	41.842	-4.6	67	0.00
63	Azobenzene	40.000	41.221	-3.1	69	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.176	-17.9	70	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.024	-2.6	70	0.00
67	4-Bromophenyl-phenylether	40.000	41.887	-4.7	71	0.00
68	Hexachlorobenzene	40.000	43.213	-8.0	73	0.00
69	Atrazine	40.000	39.918	0.2	68	0.00
70 C	Pentachlorophenol	40.000	35.207	12.0	55	0.00
71	Phenanthrene	40.000	40.716	-1.8	70	0.00
72	Anthracene	40.000	41.073	-2.7	70	0.00
73	Carbazole	40.000	39.799	0.5	68	0.00
74	Di-n-butylphthalate	40.000	40.234	-0.6	68	0.00
75 C	Fluoranthene	40.000	40.515	-1.3	69	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	66	0.00
77	Benzydine	40.000	50.778	-26.9#	78	0.00
78	Pyrene	40.000	42.102	-5.3	69	0.00
79 S	Terphenyl-d14	80.000	88.653	-10.8	76	0.00
80	Butylbenzylphthalate	40.000	43.670	-9.2	69	0.00
81	Benzo(a)anthracene	40.000	40.966	-2.4	67	0.00
82	3,3'-Dichlorobenzidine	40.000	37.393	6.5	61	0.00
83	Chrysene	40.000	40.492	-1.2	67	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.187	-5.5	68	0.00
85 c	Di-n-octyl phthalate	40.000	40.888	-2.2	64	0.00
86 I	Perylene-d12	20.000	20.000	0.0	59	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.932	-2.3	57	0.00
88	Benzo(b)fluoranthene	40.000	43.371	-8.4	64	0.00
89	Benzo(k)fluoranthene	40.000	41.445	-3.6	59	0.00
90 C	Benzo(a)pyrene	40.000	41.807	-4.5	60	0.00
91	Dibenzo(a,h)anthracene	40.000	41.223	-3.1	57	0.00
92	Benzo(g,h,i)perylene	40.000	40.315	-0.8	56	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050622\
Data File : BF128116.D
Acq On : 06 May 2022 10:08
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 06 16:06:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 17:56:23 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0