

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050625\
 Data File : BF142309.D
 Acq On : 06 May 2025 13:56
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 14:30:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 05 18:41:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.528	0.515	2.5	108	0.00
3	Pyridine	1.239	1.232	0.6	111	0.00
4	n-Nitrosodimethylamine	0.720	0.701	2.6	106	0.00
5 S	2-Fluorophenol	1.172	1.131	3.5	106	0.00
6	Aniline	1.413	1.439	-1.8	104	0.00
7 S	Phenol-d6	1.441	1.379	4.3	106	0.00
8	2-Chlorophenol	1.252	1.242	0.8	107	0.00
9	Benzaldehyde	0.843	0.692	17.9	100	0.00
10 C	Phenol	1.527	1.481	3.0	106	0.00
11	bis(2-Chloroethyl)ether	1.506	1.326	12.0	104	0.00
12	1,3-Dichlorobenzene	1.432	1.367	4.5	107	0.00
13 C	1,4-Dichlorobenzene	1.438	1.366	5.0	106	0.00
14	1,2-Dichlorobenzene	1.380	1.315	4.7	105	0.00
15	Benzyl Alcohol	0.933	0.941	-0.9	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.249	2.135	5.1	106	0.00
17	2-Methylphenol	1.014	0.982	3.2	107	0.00
18	Hexachloroethane	0.478	0.471	1.5	108	0.00
19 P	n-Nitroso-di-n-propylamine	0.891	0.864	3.0	106	0.00
20	3+4-Methylphenols	1.265	1.216	3.9	105	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.445	0.410	7.9	106	0.00
23 S	Nitrobenzene-d5	0.328	0.334	-1.8	109	0.00
24	Nitrobenzene	0.309	0.312	-1.0	109	0.00
25	Isophorone	0.621	0.590	5.0	108	0.00
26 C	2-Nitrophenol	0.129	0.144	-11.6	117	0.00
27	2,4-Dimethylphenol	0.312	0.299	4.2	107	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.375	6.0	107	0.00
29 C	2,4-Dichlorophenol	0.265	0.262	1.1	107	0.00
30	1,2,4-Trichlorobenzene	0.299	0.282	5.7	107	0.00
31	Naphthalene	0.969	0.899	7.2	106	0.00
32	Benzoic acid	0.150	0.160	-6.7	123	0.00
33	4-Chloroaniline	0.396	0.371	6.3	104	0.00
34 C	Hexachlorobutadiene	0.177	0.168	5.1	107	0.00
35	Caprolactam	0.077	0.082	-6.5	112	0.00
36 C	4-Chloro-3-methylphenol	0.276	0.264	4.3	107	0.00
37	2-Methylnaphthalene	0.602	0.560	7.0	105	0.00
38	1-Methylnaphthalene	0.628	0.581	7.5	105	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.559	0.522	6.6	107	0.00
41 P	Hexachlorocyclopentadiene	0.345	0.347	-0.6	107	0.00
42 S	2,4,6-Tribromophenol	0.193	0.197	-2.1	109	0.00
43 C	2,4,6-Trichlorophenol	0.352	0.357	-1.4	107	0.00
44	2,4,5-Trichlorophenol	0.375	0.374	0.3	109	0.00
45 S	2-Fluorobiphenyl	1.403	1.255	10.5	104	0.00
46	1,1'-Biphenyl	1.533	1.431	6.7	107	0.00
47	2-Chloronaphthalene	1.140	1.056	7.4	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050625\
 Data File : BF142309.D
 Acq On : 06 May 2025 13:56
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 14:30:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 05 18:41:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.292	0.322	-10.3	114	0.00
49	Acenaphthylene	1.908	1.782	6.6	106	0.00
50	Dimethylphthalate	1.284	1.233	4.0	108	0.00
51	2,6-Dinitrotoluene	0.240	0.261	-8.8	112	0.00
52 C	Acenaphthene	1.147	1.069	6.8	106	0.00
53	3-Nitroaniline	0.282	0.297	-5.3	110	0.00
54 P	2,4-Dinitrophenol	0.102	0.109	-6.9	119	0.00
55	Dibenzofuran	1.693	1.570	7.3	107	0.00
56 P	4-Nitrophenol	0.212	0.228	-7.5	113	0.00
57	2,4-Dinitrotoluene	0.308	0.340	-10.4	113	0.00
58	Fluorene	1.292	1.189	8.0	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.309	0.321	-3.9	111	0.00
60	Diethylphthalate	1.283	1.230	4.1	108	0.00
61	4-Chlorophenyl-phenylether	0.623	0.577	7.4	106	0.00
62	4-Nitroaniline	0.223	0.219	1.8	103	0.00
63	Azobenzene	1.226	1.147	6.4	107	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	0.082	0.087	-6.1	118	0.00
66 c	n-Nitrosodiphenylamine	0.666	0.618	7.2	107	0.00
67	4-Bromophenyl-phenylether	0.227	0.213	6.2	108	0.00
68	Hexachlorobenzene	0.249	0.235	5.6	109	0.00
69	Atrazine	0.175	0.186	-6.3	116	0.00
70 C	Pentachlorophenol	0.131	0.140	-6.9	110	0.00
71	Phenanthrene	1.048	0.966	7.8	107	0.00
72	Anthracene	1.077	0.996	7.5	107	0.00
73	Carbazole	0.966	0.916	5.2	108	0.00
74	Di-n-butylphthalate	1.016	1.042	-2.6	113	0.00
75 C	Fluoranthene	1.048	1.012	3.4	112	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
77	Benzidine	0.363	0.404	-11.3	127	0.00
78	Pyrene	1.780	1.651	7.2	109	0.00
79 S	Terphenyl-d14	1.351	1.241	8.1	112	0.00
80	Butylbenzylphthalate	0.449	0.529	-17.8	127	0.00
81	Benzo(a)anthracene	1.295	1.235	4.6	113	0.00
82	3,3'-Dichlorobenzidine	0.301	0.309	-2.7	118	0.00
83	Chrysene	1.206	1.120	7.1	114	0.00
84	Bis(2-ethylhexyl)phthalate	0.573	0.593	-3.5	111	0.00
85 c	Di-n-octyl phthalate	1.057	0.934	11.6	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.391	1.368	1.7	101	0.00
88	Benzo(b)fluoranthene	1.211	1.188	1.9	104	0.00
89	Benzo(k)fluoranthene	1.109	1.039	6.3	102	0.00
90 C	Benzo(a)pyrene	1.085	1.061	2.2	101	0.00
91	Dibenzo(a,h)anthracene	1.149	1.114	3.0	100	0.00
92	Benzo(g,h,i)perylene	1.141	1.112	2.5	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050625\
Data File : BF142309.D
Acq On : 06 May 2025 13:56
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 06 14:30:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 05 18:41:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0