

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF050716\  
 Data File : BF086782.D  
 Acq On : 7 May 2016 19:06  
 Operator : UM/SJ  
 Sample : H2878-03  
 Misc :  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 URSMW-08

Quant Time: May 08 00:25:34 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF050416.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat May 07 23:59:39 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.70  | 152  | 172522   | 40.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.98  | 136  | 687894   | 40.00  | ng    | -0.01    |
| 38) Acenaphthene-d10        | 9.73  | 164  | 330461   | 40.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.22 | 188  | 611754   | 40.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.85 | 240  | 400043   | 40.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.22 | 264  | 312695   | 40.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.31  | 112  | 854017   | 81.95  | ng    | 0.01     |
| 7) Phenol-d6                | 6.36  | 99   | 745781   | 53.13  | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.28  | 82   | 1354435  | 101.18 | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.53 | 330  | 566421   | 175.22 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.07  | 172  | 2178574  | 108.07 | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.81 | 244  | 1809511  | 91.42  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF050716\  
Data File : BF086782.D  
Acq On : 7 May 2016 19:06  
Operator : UM/SJ  
Sample : H2878-03  
Misc :  
ALS Vial : 52 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
URSMW-08

Quant Time: May 08 00:25:34 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF050416.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat May 07 23:59:39 2016  
Response via : Initial Calibration

