

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050716\
 Data File : BF086813.D
 Acq On : 8 May 2016 10:05
 Operator : UM/SJ
 Sample : H2762-08DL 20X
 Misc :
 ALS Vial : 81 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-B4COMPDL

Manual Integrations
 APPROVED

Sohil
 5/9/2016 7:20:36 PM

Quant Time: May 09 18:35:37 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:12:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	181780	40.00	ng	-0.03
21) Naphthalene-d8	8.00	136	707659	40.00	ng	-0.03
38) Acenaphthene-d10	9.73	164	292531	40.00	ng	-0.05
63) Phenanthrene-d10	11.22	188	488256	40.00	ng	-0.03
75) Chrysene-d12	13.85	240	352889	40.00	ng	-0.03
86) Perylene-d12	15.22	264	310917	40.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	74426	6.78	ng	-0.01
7) Phenol-d6	6.36	99	108243	7.32	ng	-0.01
23) Nitrobenzene-d5	7.28	82	60202	4.37	ng	-0.03
41) 2,4,6-Tribromophenol	10.53	330	17197	6.01	ng	-0.03
44) 2-Fluorobiphenyl	9.07	172	105577	5.92	ng	-0.03
78) Terphenyl-d14	12.80	244	72161	4.13	ng	-0.05
Target Compounds						Qvalue
31) Naphthalene	8.02	128	1656215	45.81	ng	100
37) 2-Methylnaphthalene	8.70	142	525377	24.63	ng	99
45) 1,1'-Biphenyl	9.16	154	100529	4.20	ng	95
48) Acenaphthylene	9.60	152	284878	10.40	ng	97
51) Acenaphthene	9.77	154	117511	6.50	ng	97
57) Fluorene	10.28	166	281740	15.52	ng	99
70) Phenanthrene	11.24	178	900193	34.32	ng	99
71) Anthracene	11.29	178	269511	9.81	ng	99
74) Fluoranthene	12.42	202	272939	10.38	ng	99
77) Pyrene	12.65	202	355522	12.64	ng	99
80) Benzo(a)anthracene	13.84	228	105081m	5.27	ng	
82) Chrysene	13.87	228	95679m	4.49	ng	
89) Benzo(a)pyrene	15.17	252	62188	3.61	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050716\
Data File : BF086813.D
Acq On : 8 May 2016 10:05
Operator : UM/SJ
Sample : H2762-08DL 20X
Misc :
ALS Vial : 81 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-B4COMPDL

Manual Integrations
APPROVED

Sohil
5/9/2016 7:20:36 PM

Quant Time: May 09 18:35:37 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 15:12:53 2016
Response via : Initial Calibration

