

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
 Data File : BF079038.D
 Acq On : 8 May 2015 15:24
 Operator : TP/IZ
 Sample : PB83243BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB83243BL

Quant Time: May 08 23:01:30 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 08 22:44:53 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.32	152	68203	20.00	ng	0.00
21) Naphthalene-d8	8.90	136	287234	20.00	ng	0.00
38) Acenaphthene-d10	11.07	164	132716	20.00	ng	0.00
63) Phenanthrene-d10	12.93	188	258758	20.00	ng	0.00
75) Chrysene-d12	16.26	240	260340	20.00	ng	0.01
86) Perylene-d12	18.15	264	254362	20.00	ng	0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	577231	145.69	ng	0.01
7) Phenol-d6	6.89	99	733693	140.09	ng	0.00
23) Nitrobenzene-d5	8.01	82	399296	79.89	ng	0.00
41) 2,4,6-Tribromophenol	12.06	330	197113	137.29	ng	0.00
44) 2-Fluorobiphenyl	10.25	172	828514	86.97	ng	0.00
78) Terphenyl-d14	14.93	244	968322	89.05	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
Data File : BF079038.D
Acq On : 8 May 2015 15:24
Operator : TP/IZ
Sample : PB83243BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB83243BL

Quant Time: May 08 23:01:30 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 08 22:44:53 2015
Response via : Initial Calibration

