

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
 Data File : BF079087.D
 Acq On : 9 May 2015 18:31
 Operator : TP/IZ
 Sample : G2035-15RE
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-14-D3RE

Quant Time: May 11 02:07:42 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 11 01:50:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.32	152	177925	20.00	ng	0.00
21) Naphthalene-d8	8.89	136	749355	20.00	ng	0.00
38) Acenaphthene-d10	11.07	164	392435	20.00	ng	0.00
63) Phenanthrene-d10	12.91	188	710381	20.00	ng	0.00
75) Chrysene-d12	16.21	240	667902	20.00	ng	0.00
86) Perylene-d12	17.94	264	677633	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	786833	76.12	ng	0.00
7) Phenol-d6	6.88	99	980186	71.74	ng	0.00
23) Nitrobenzene-d5	8.01	82	596360	45.74	ng	0.00
41) 2,4,6-Tribromophenol	12.06	330	298187	70.24	ng	0.00
44) 2-Fluorobiphenyl	10.24	172	953005	33.83	ng	0.00
78) Terphenyl-d14	14.91	244	1016692	36.44	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.74	163	58644	2.04	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
Data File : BF079087.D
Acq On : 9 May 2015 18:31
Operator : TP/IZ
Sample : G2035-15RE
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-14-D3RE

Quant Time: May 11 02:07:42 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 11 01:50:36 2015
Response via : Initial Calibration

