

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
 Data File : BF079089.D
 Acq On : 9 May 2015 19:28
 Operator : TP/IZ
 Sample : G2115-01RE
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUMBO-EP-5RE

Quant Time: May 11 02:25:35 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 11 01:50:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.31	152	86859	20.00	ng	-0.01
21) Naphthalene-d8	8.89	136	364525	20.00	ng	0.00
38) Acenaphthene-d10	11.07	164	181594	20.00	ng	0.00
63) Phenanthrene-d10	12.91	188	340703	20.00	ng	0.00
75) Chrysene-d12	16.22	240	323933	20.00	ng	0.01
86) Perylene-d12	17.97	264	337597	20.00	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	188673	37.39	ng	0.00
7) Phenol-d6	6.88	99	310569	46.56	ng	0.00
23) Nitrobenzene-d5	8.01	82	180169	28.41	ng	0.00
41) 2,4,6-Tribromophenol	12.06	330	145774	74.20	ng	0.00
44) 2-Fluorobiphenyl	10.24	172	537795	41.26	ng	0.00
78) Terphenyl-d14	14.91	244	712098	52.63	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.74	163	36273	2.72	ng	Qvalue 100
74) Fluoranthene	14.42	202	54587	2.75	ng	95
77) Pyrene	14.71	202	49048	2.63	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050815\
Data File : BF079089.D
Acq On : 9 May 2015 19:28
Operator : TP/IZ
Sample : G2115-01RE
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DUMBO-EP-5RE

Quant Time: May 11 02:25:35 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 11 01:50:36 2015
Response via : Initial Calibration

