


Data Path : Z:\HPCHEM1\BNA F\Data\BF050817\
 Data File : BF094982.D
 Acq On : 8 May 2017 16:11
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2307, 2308, 2309; Background Corrected with Scan 2301

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.2	16636	PASS
68	69	0.00	2	1.8	289	PASS
69	198	0.00	100	34.5	16317	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.3	22367	PASS
197	198	0.00	2	0.7	312	PASS
198	198	100	100	100.0	47240	PASS
199	198	5	9	6.8	3227	PASS
275	198	10	60	24.4	11503	PASS
365	198	1	100	2.4	1148	PASS
441	442	0.01	24	16.4	6629	PASS
442	198	50	100	85.4	40365	PASS
443	442	15	24	19.5	7858	PASS