

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050916\
 Data File : BF086833.D
 Acq On : 9 May 2016 21:07
 Operator : UM/SJ
 Sample : PB90374BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB90374BL

Quant Time: May 10 03:22:35 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 16:04:56 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	151563	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	631576	20.00	ng	-0.01
38) Acenaphthene-d10	9.72	164	316443	20.00	ng	-0.01
63) Phenanthrene-d10	11.21	188	663020	20.00	ng	0.00
75) Chrysene-d12	13.83	240	567452	20.00	ng	-0.01
86) Perylene-d12	15.21	264	471008	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	1169679	130.70	ng	0.01
7) Phenol-d6	6.35	99	1430599	119.61	ng	-0.01
23) Nitrobenzene-d5	7.26	82	1026239	81.59	ng	0.00
41) 2,4,6-Tribromophenol	10.52	330	439411	131.16	ng	0.00
44) 2-Fluorobiphenyl	9.06	172	1686806	89.59	ng	0.00
78) Terphenyl-d14	12.78	244	1741615	65.12	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050916\
Data File : BF086833.D
Acq On : 9 May 2016 21:07
Operator : UM/SJ
Sample : PB90374BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB90374BL

Quant Time: May 10 03:22:35 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 16:04:56 2016
Response via : Initial Calibration

