


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050916\
 Data File : BF086816.D
 Acq On : 9 May 2016 10:14
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050916.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 09 16:04:56 2016



AutoFind: Scans 820, 821, 822; Background Corrected with Scan 817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	63.7	57883	PASS
68	69	0.00	2	1.9	757	PASS
69	198	0.00	100	44.7	40662	PASS
70	69	0.00	2	0.6	224	PASS
127	198	10	80	50.9	46251	PASS
197	198	0.00	2	0.9	775	PASS
198	198	100	100	100.0	90928	PASS
199	198	5	9	6.6	6029	PASS
275	198	10	60	26.8	24371	PASS
365	198	1	100	3.9	3585	PASS
441	442	0.01	24	15.8	13736	PASS
442	198	50	100	95.7	87026	PASS
443	442	15	24	19.0	16529	PASS