


Data Path : Z:\HPCHEM1\BNA_F\Data\BF050917\
 Data File : BF095047.D
 Acq On : 10 May 2017 5:44
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2306, 2307, 2308; Background Corrected with Scan 2301

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	15111	PASS
68	69	0.00	2	0.8	126	PASS
69	198	0.00	100	35.3	15221	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.6	20111	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	43152	PASS
199	198	5	9	7.2	3089	PASS
275	198	10	60	21.1	9111	PASS
365	198	1	100	2.2	941	PASS
441	442	0.01	24	15.1	5191	PASS
442	198	50	100	79.7	34410	PASS
443	442	15	24	20.1	6908	PASS