

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050922\
 Data File : BF128141.D
 Acq On : 09 May 2022 11:31
 Operator : CG\JU
 Sample : SSTDICC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC001

Quant Time: May 09 14:08:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 14:06:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	96955	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	379751	20.000	ng	0.00
39) Acenaphthene-d10	9.934	164	221485	20.000	ng	0.00
64) Phenanthrene-d10	11.428	188	422276	20.000	ng	0.00
76) Chrysene-d12	14.074	240	369931	20.000	ng	0.00
86) Perylene-d12	15.574	264	345582	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.293	70	5408	1.119	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050922\
Data File : BF128141.D
Acq On : 09 May 2022 11:31
Operator : CG\JU
Sample : SSTDICC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC001

Quant Time: May 09 14:08:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 14:06:17 2022
Response via : Initial Calibration

