


Benzaldehyde


Response Ratio



$R = -1.554e-001 A^2 + 1.118e+000 A + 6.628e-002$
Coef of Det (r^2) = 0.997604 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF050922.M
Calibration Table Last Updated: Mon May 09 15:30:11 2022

Hexachlorocyclopentadiene

Response Ratio



$$\text{Response} = 2.881\text{e-}001 * \text{Amt} - 7.384\text{e-}002$$

Coef of Det (r^2) = 0.998359 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF050922.M

Calibration Table Last Updated: Mon May 09 15:30:11 2022