

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051016\
 Data File : BF086880.D
 Acq On : 11 May 2016 4:43
 Operator : UM/SJ
 Sample : H2768-08DL 5X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SA(1)8DL

Manual Integrations
 APPROVED

sohil
 5/11/2016 8:17:59 PM

Quant Time: May 11 15:14:10 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 16:04:56 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	138269	20.00	ng	-0.01
21) Naphthalene-d8	7.97	136	485888	20.00	ng	-0.01
38) Acenaphthene-d10	9.72	164	194805	20.00	ng	-0.01
63) Phenanthrene-d10	11.20	188	295572	20.00	ng	-0.01
75) Chrysene-d12	13.83	240	266872	20.00	ng	-0.01
86) Perylene-d12	15.20	264	229726	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	163100	19.98	ng	-0.01
7) Phenol-d6	6.34	99	222612	20.40	ng	-0.02
23) Nitrobenzene-d5	7.25	82	161185	16.66	ng	-0.01
41) 2,4,6-Tribromophenol	10.51	330	29881	14.49	ng	-0.01
44) 2-Fluorobiphenyl	9.05	172	231791	20.00	ng	-0.01
78) Terphenyl-d14	12.77	244	163032	12.96	ng	-0.02
Target Compounds						Qvalue
31) Naphthalene	8.00	128	222848m	9.16	ng	98
37) 2-Methylnaphthalene	8.69	142	544551	38.27	ng	98
45) 1,1'-Biphenyl	9.15	154	103345	6.67	ng	98
51) Acenaphthene	9.76	154	39737	3.58	ng	# 90
57) Fluorene	10.27	166	76382	6.28	ng	# 86
70) Phenanthrene	11.22	178	201990	12.81	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051016\
Data File : BF086880.D
Acq On : 11 May 2016 4:43
Operator : UM/SJ
Sample : H2768-08DL 5X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SA(1)8DL

Manual Integrations
APPROVED

sohil
5/11/2016 8:17:59 PM

Quant Time: May 11 15:14:10 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 16:04:56 2016
Response via : Initial Calibration

