


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051017\
 Data File : BF095055.D
 Acq On : 10 May 2017 14:20
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2306, 2307, 2308; Background Corrected with Scan 2300

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.6	13530	PASS
68	69	0.00	2	1.0	139	PASS
69	198	0.00	100	34.0	13714	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.0	19731	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	40288	PASS
199	198	5	9	6.7	2696	PASS
275	198	10	60	23.0	9280	PASS
365	198	1	100	2.5	1021	PASS
441	442	0.01	24	15.5	5644	PASS
442	198	50	100	90.4	36426	PASS
443	442	15	24	20.0	7288	PASS