

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051021\
 Data File : BF123898.D
 Acq On : 10 May 2021 22:39
 Operator : JU/CG
 Sample : M2322-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-20210505

Quant Time: May 11 03:38:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 10 16:46:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.769	152	113333	20.00	ng	0.00
21) Naphthalene-d8	8.051	136	479437	20.00	ng	0.00
39) Acenaphthene-d10	9.804	164	231874	20.00	ng	0.00
64) Phenanthrene-d10	11.292	188	389037	20.00	ng	# 0.00
76) Chrysene-d12	13.927	240	242621	20.00	ng	# 0.00
86) Perylene-d12	15.368	264	224798	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.404	112	781694	100.73	ng	0.01
7) Phenol-d6	6.416	99	901637	87.05	ng	0.00
23) Nitrobenzene-d5	7.333	82	965568	83.60	ng	0.00
42) 2,4,6-Tribromophenol	10.598	330	386402	113.13	ng	0.00
45) 2-Fluorobiphenyl	9.133	172	1331332	71.65	ng	0.00
79) Terphenyl-d14	12.880	244	1599934	107.77	ng	0.00
Target Compounds						
60) Diethylphthalate	10.227	149	89243	4.92	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051021\
Data File : BF123898.D
Acq On : 10 May 2021 22:39
Operator : JU/CG
Sample : M2322-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-20210505

Quant Time: May 11 03:38:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 10 16:46:01 2021
Response via : Initial Calibration

