


Indeno(1,2,3-cd)pyrene


Response Ratio



$R = 2.543e-001 A^2 + 7.123e-001 A + 2.033e-001$
 Coef of Det (r^2) = 0.990463 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051021.M
 Calibration Table Last Updated: Mon May 10 16:46:01 2021

1,4-Dioxane


Response Ratio



$R = -1.840e-002 A^2 + 5.799e-001 A + 1.183e-001$
 Coef of Det (r^2) = 0.993174 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051021.M
 Calibration Table Last Updated: Mon May 10 16:46:01 2021

Hexachlorocyclopentadiene

Response Ratio



$R = 6.867e-002 A^2 + 1.699e-002 A + 6.068e-002$
Coef of Det (r^2) = 0.990984 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051021.M
Calibration Table Last Updated: Mon May 10 16:46:01 2021