

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051021\
 Data File : BF123883.D
 Acq On : 10 May 2021 13:01
 Operator : JU/CG
 Sample : SSTDIC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC010

Manual Integrations
 APPROVED

mohammad
 5/11/2021 4:15:59 PM

Quant Time: May 10 14:58:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 10 14:57:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.769	152	93764	20.00	ng	0.00
21) Naphthalene-d8	8.051	136	339789	20.00	ng	0.00
39) Acenaphthene-d10	9.810	164	161935	20.00	ng	0.00
64) Phenanthrene-d10	11.292	188	320159	20.00	ng	0.00
76) Chrysene-d12	13.933	240	266479	20.00	ng	# 0.00
86) Perylene-d12	15.368	264	185035	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	174398	29.47	ng	0.00
7) Phenol-d6	6.410	99	192620	23.63	ng	0.00
23) Nitrobenzene-d5	7.334	82	189433	25.17	ng	0.00
42) 2,4,6-Tribromophenol	10.592	330	52820	26.20	ng	0.00
45) 2-Fluorobiphenyl	9.128	172	287837	25.89	ng	0.00
79) Terphenyl-d14	12.874	244	326947	23.63	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.463	88	41023	12.87	ng	# 93
3) Pyridine	3.204	79	94856	12.17	ng	99
4) n-Nitrosodimethylamine	3.128	42	69073	16.09	ng	# 83
6) Aniline	6.428	93	103683	11.00	ng	# 59
8) 2-Chlorophenol	6.557	128	75864	11.92	ng	95
9) Benzaldehyde	6.316	77	46230	11.72	ng	97
10) Phenol	6.422	94	101340	11.57	ng	# 72
11) bis(2-Chloroethyl)ether	6.504	93	69625	10.89	ng	98
12) 1,3-Dichlorobenzene	6.710	146	77302	12.02	ng	99
13) 1,4-Dichlorobenzene	6.787	146	78884	12.12	ng	94
14) 1,2-Dichlorobenzene	6.939	146	77042	12.19	ng	96
15) Benzyl Alcohol	6.910	79	78609m	12.30	ng	
16) 2,2'-oxybis(1-Chloropr...	7.045	45	151481	11.14	ng	99
17) 2-Methylphenol	7.028	107	64051	11.81	ng	# 93
18) Hexachloroethane	7.281	117	31461	12.21	ng	97
19) n-Nitroso-di-n-propyla...	7.181	70	56796	11.28	ng	# 81
20) 3+4-Methylphenols	7.187	107	82850	12.00	ng	# 84
22) Acetophenone	7.175	105	112618	12.19	ng	# 91
24) Nitrobenzene	7.351	77	96304	12.28	ng	97
25) Isophorone	7.592	82	155778	11.04	ng	96
26) 2-Nitrophenol	7.669	139	36309	11.35	ng	# 81
27) 2,4-Dimethylphenol	7.710	122	55204	10.98	ng	91
28) bis(2-Chloroethoxy)met...	7.804	93	75241	10.46	ng	98
29) 2,4-Dichlorophenol	7.916	162	56730	11.45	ng	98
30) 1,2,4-Trichlorobenzene	7.992	180	61272	11.80	ng	98
31) Naphthalene	8.075	128	208481	11.91	ng	98
32) Benzoic acid	7.810	122	27164	8.59	ng	89
33) 4-Chloroaniline	8.128	127	84211	11.53	ng	98
34) Hexachlorobutadiene	8.198	225	41213	13.03	ng	98
35) Caprolactam	8.469	113	18525	10.65	ng	# 83
36) 4-Chloro-3-methylphenol	8.610	107	72684	11.76	ng	97
37) 2-Methylnaphthalene	8.769	142	132053	11.58	ng	99
38) 1-Methylnaphthalene	8.863	142	128514	11.99	ng	97
40) 1,2,4,5-Tetrachloroben...	8.933	216	61184	12.84	ng	98
41) Hexachlorocyclopentadiene	8.922	237	10722	5.67	ng	94
43) 2,4,6-Trichlorophenol	9.051	196	39227	11.95	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051021\
 Data File : BF123883.D
 Acq On : 10 May 2021 13:01
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Manual Integrations
 APPROVED

mohammad
 5/11/2021 4:15:59 PM


Quant Time: May 10 14:58:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 10 14:57:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.092	196	40270	11.43	ng	95
46) 1,1'-Biphenyl	9.228	154	158748	11.97	ng	100
47) 2-Chloronaphthalene	9.257	162	124031	12.40	ng	96
48) 2-Nitroaniline	9.351	65	52515	11.92	ng	89
49) Acenaphthylene	9.669	152	199821	12.38	ng	98
50) Dimethylphthalate	9.528	163	133779	11.73	ng	99
51) 2,6-Dinitrotoluene	9.592	165	28887	10.95	ng	84
52) Acenaphthene	9.839	154	120500	12.25	ng	97
53) 3-Nitroaniline	9.763	138	35295	11.21	ng	94
54) 2,4-Dinitrophenol	9.875	184	11709	8.36	ng	# 87
55) Dibenzofuran	10.010	168	180827	12.15	ng	94
56) 4-Nitrophenol	9.933	139	21138	9.22	ng	# 82
57) 2,4-Dinitrotoluene	9.992	165	39824	11.08	ng	# 76
58) Fluorene	10.357	166	140434	12.28	ng	97
59) 2,3,4,6-Tetrachlorophenol	10.133	232	32870	11.79	ng	99
60) Diethylphthalate	10.228	149	141535	11.66	ng	97
61) 4-Chlorophenyl-phenyle...	10.351	204	67054	12.58	ng	94
62) 4-Nitroaniline	10.369	138	34697	11.12	ng	# 83
63) Azobenzene	10.510	77	161498	11.57	ng	95
65) 4,6-Dinitro-2-methylph...	10.404	198	19456	9.81	ng	99
66) n-Nitrosodiphenylamine	10.463	169	123138	11.76	ng	96
67) 4-Bromophenyl-phenylether	10.833	248	41398	12.28	ng	# 91
68) Hexachlorobenzene	10.904	284	47933	12.49	ng	# 93
69) Atrazine	10.992	200	35266	11.14	ng	96
70) Pentachlorophenol	11.104	266	17020	8.70	ng	90
71) Phenanthrene	11.316	178	205048	12.21	ng	97
72) Anthracene	11.369	178	199644	11.96	ng	99
73) Carbazole	11.522	167	208615	12.36	ng	99
74) Di-n-butylphthalate	11.857	149	201052	10.61	ng	98
75) Fluoranthene	12.504	202	219235	11.97	ng	95
77) Benzidine	12.627	184	50373	7.30	ng	99
78) Pyrene	12.733	202	231614	11.20	ng	98
80) Butylbenzylphthalate	13.351	149	82706	9.36	ng	96
81) Benzo(a)anthracene	13.921	228	192042	11.89	ng	98
82) 3,3'-Dichlorobenzidine	13.880	252	54906	11.07	ng	98
83) Chrysene	13.957	228	187849	11.56	ng	99
84) Bis(2-ethylhexyl)phtha...	13.916	149	97291	8.86	ng	96
85) Di-n-octyl phthalate	14.527	149	136705	7.79	ng	100
87) Indeno(1,2,3-cd)pyrene	16.786	276	118481	10.99	ng	# 94
88) Benzo(b)fluoranthene	14.951	252	150825	12.10	ng	97
89) Benzo(k)fluoranthene	14.980	252	139265m	11.70	ng	
90) Benzo(a)pyrene	15.310	252	123795	11.61	ng	# 94
91) Dibenzo(a,h)anthracene	16.798	278	99002	10.88	ng	# 97
92) Benzo(g,h,i)perylene	17.209	276	101814	11.17	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

mohammad
5/11/2021 4:15:59 PM




[illegible]




#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.769 min Scan# 796
Delta R.T. -0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion: 152 Resp: 93764
Ion Ratio Lower Upper
152 100
150 163.1 127.0 190.6
115 69.3 54.3 81.5



#2
1,4-Dioxane
Concen: 12.87 ng
RT: 2.463 min Scan# 64
Delta R.T. 0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion: 88 Resp: 41023
Ion Ratio Lower Upper
88 100
58 110.1 83.8 125.8
43 48.4 32.2 48.4#



Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#3
Pyridine
Concen: 12.17 ng
RT: 3.204 min Scan# 190
Delta R.T. 0.024 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion: 79 Resp: 94856
Ion Ratio Lower Upper
79 100
52 108.9 87.4 131.0
51 49.5 41.4 62.0

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#4
n-Nitrosodimethylamine
Concen: 16.09 ng
RT: 3.128 min Scan# 177
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion: 42 Resp: 69073
Ion Ratio Lower Upper
42 100
74 86.1 83.5 125.3
44 9.4 5.7 8.5









#5
2-Fluorophenol
Concen: 29.47 ng
RT: 5.387 min Scan# 561
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 112 Resp: 174398
Ion Ratio Lower Upper
112 100
64 82.7 61.4 92.0
63 46.0 33.0 49.4

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#6
Aniline
Concen: 11.00 ng
RT: 6.428 min Scan# 738
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 93 Resp: 103683
Ion Ratio Lower Upper
93 100
66 74.1 36.4 54.6#
65 42.6 19.4 29.0#









#7
Phenol-d6
Concen: 23.63 ng
RT: 6.410 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 99 Resp: 192620
Ion Ratio Lower Upper
99 100
42 26.9 19.8 29.8
71 45.2 31.1 46.7

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#8
2-Chlorophenol
Concen: 11.92 ng
RT: 6.557 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 128 Resp: 75864
Ion Ratio Lower Upper
128 100
130 28.5 12.6 52.6
64 58.3 41.5 81.5









#9
Benzaldehyde
Concen: 11.72 ng
RT: 6.316 min Scan# 719
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:	77	Resp:	46230
Ion Ratio	Lower	Upper	
77	100		
105	95.4	75.2	115.2
106	85.4	71.6	111.6

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#10
Phenol
Concen: 11.57 ng
RT: 6.422 min Scan# 737
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:	94	Resp:	101340
Ion Ratio	Lower	Upper	
94	100		
65	49.0	18.1	58.1
66	73.6	29.0	69.0









#11
bis(2-Chloroethyl)ether
Concen: 10.89 ng
RT: 6.504 min Scan# 751
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 93 Resp: 69625
Ion Ratio Lower Upper
93 100
63 97.6 76.1 116.1
95 34.5 11.5 51.5

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#12
1,3-Dichlorobenzene
Concen: 12.02 ng
RT: 6.710 min Scan# 786
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 146 Resp: 77302
Ion Ratio Lower Upper
146 100
148 63.1 50.9 76.3
75 44.6 34.2 51.4





#13

1,4-Dichlorobenzene

Concen: 12.12 ng

RT: 6.787 min Scan# 799

Delta R.T. 0.000 min

Lab File: BF123883.D

Acq: 10 May 2021 13:01

Tgt Ion:	146	Resp:	78884
Ion	Ratio	Lower	Upper
146	100		
148	68.9	49.9	74.9
111	47.9	39.5	59.3

Instrument :





BNA_F

ClientSampled :

SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#14

1,2-Dichlorobenzene

Concen: 12.19 ng




RT: 6.939 min Scan# 825


Delta R.T. -0.006 min

Lab File: BF123883.D

Acq: 10 May 2021 13:01


Tgt Ion:	146	Resp:	77042
Ion	Ratio	Lower	Upper
146	100		
148	61.9	51.6	77.4
111	51.6	38.4	57.6





#15
Benzyl Alcohol
Concen: 12.30 ng m
RT: 6.910 min Scan# 820
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion: 79 Resp: 78609
Ion Ratio Lower Upper
79 100
108 64.1 55.8 83.8
77 65.5 49.9 74.9

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 11.14 ng
RT: 7.045 min Scan# 843
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion: 45 Resp: 151481
Ion Ratio Lower Upper
45 100
77 10.4 0.0 30.0
79 7.6 0.0 27.3









#17
2-Methylphenol
Concen: 11.81 ng
RT: 7.028 min Scan# 840
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:	107	Resp:	64051
Ion	Ratio	Lower	Upper
107	100		
108	109.8	89.8	134.8
77	61.5	40.9	61.3#
79	57.0	40.2	60.4

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#18
Hexachloroethane
Concen: 12.21 ng
RT: 7.281 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:	117	Resp:	31461
Ion	Ratio	Lower	Upper
117	100		
119	96.2	77.8	116.8
201	89.5	67.8	101.8








#19
n-Nitroso-di-n-propylamine
Concen: 11.28 ng
RT: 7.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:	70	Resp:	56796
Ion	Ratio	Lower	Upper
70	100		
42	93.9	60.1	90.1#
101	9.9	7.4	11.2
130	23.6	14.2	21.2#



#20
3+4-Methylphenols
Concen: 12.00 ng
RT: 7.187 min Scan# 867
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01


Tgt Ion:	107	Resp:	82850
Ion	Ratio	Lower	Upper
107	100		
108	96.2	71.5	111.5
77	74.1	82.0	122.0#
79	38.6	14.1	54.1



Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.051 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Instrument :
BNA_F
ClientSampleId :
SSTDIC010






Tgt Ion:136 Resp: 339789


Ion	Ratio	Lower	Upper
136	100		
137	11.9	8.7	13.1
54	13.0	9.8	14.6
68	5.3	4.1	6.1

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM






#22
Acetophenone
Concen: 12.19 ng
RT: 7.175 min Scan# 865
Delta R.T. -0.012 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:105 Resp: 112618

Ion	Ratio	Lower	Upper
105	100		
71	6.8	4.2	6.4#
51	51.0	33.6	50.4#
120	20.5	16.3	24.5









#23
Nitrobenzene-d5
Concen: 25.17 ng
RT: 7.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 82 Resp: 189433
Ion Ratio Lower Upper
82 100
128 36.2 29.4 44.2
54 71.6 54.1 81.1

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#24
Nitrobenzene
Concen: 12.28 ng
RT: 7.351 min Scan# 895
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 77 Resp: 96304
Ion Ratio Lower Upper
77 100
123 40.9 34.5 51.7
65 14.7 12.2 18.2









#25
Isophorone
Concen: 11.04 ng
RT: 7.592 min Scan# 936
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 82 Resp: 155778
Ion Ratio Lower Upper
82 100
95 8.4 5.8 8.6
138 16.6 11.8 17.8

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#26
2-Nitrophenol
Concen: 11.35 ng
RT: 7.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 139 Resp: 36309
Ion Ratio Lower Upper
139 100
109 46.7 28.6 43.0#
65 87.4 56.8 85.2#





#27

2,4-Dimethylphenol

Concen: 10.98 ng

RT: 7.710 min Scan# 956

Delta R.T. -0.006 min

Lab File: BF123883.D

Acq: 10 May 2021 13:01

Tgt Ion:	122	Resp:	55204
Ion	Ratio	Lower	Upper
122	100		
107	147.4	106.3	159.5
121	58.7	45.4	68.0

Instrument :





BNA_F

ClientSampleId :

SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#28

bis(2-Chloroethoxy)methane

Concen: 10.46 ng




RT: 7.804 min Scan# 972


Delta R.T. 0.000 min

Lab File: BF123883.D

Acq: 10 May 2021 13:01


Tgt Ion:	93	Resp:	75241
Ion	Ratio	Lower	Upper
93	100		
95	34.4	26.5	39.7
123	9.8	8.4	12.6





#29
2,4-Dichlorophenol
Concen: 11.45 ng
RT: 7.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:162 Resp: 56730
Ion Ratio Lower Upper
162 100
164 66.3 44.1 84.1
98 47.2 27.3 67.3

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#30
1,2,4-Trichlorobenzene
Concen: 11.80 ng
RT: 7.992 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:180 Resp: 61272
Ion Ratio Lower Upper
180 100
182 95.0 77.2 115.8
145 34.9 26.0 39.0









#31
Naphthalene
Concen: 11.91 ng
RT: 8.075 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:128 Resp: 208481
Ion Ratio Lower Upper
128 100
129 10.6 8.9 13.3
127 12.5 10.8 16.2

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM




#32
Benzoic acid
Concen: 8.59 ng
RT: 7.810 min Scan# 973
Delta R.T. -0.035 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:122 Resp: 27164
Ion Ratio Lower Upper
122 100
105 141.6 107.8 147.8
77 111.4 81.6 121.6









#33
4-Chloroaniline
Concen: 11.53 ng
RT: 8.128 min Scan# 1027
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:	127	Resp:	84211
Ion Ratio	Lower	Upper	
127	100		
129	30.5	25.7	38.5
65	42.9	34.5	51.7
92	20.0	17.4	26.0



#34
Hexachlorobutadiene
Concen: 13.03 ng
RT: 8.198 min Scan# 1039
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:	225	Resp:	41213
Ion Ratio	Lower	Upper	
225	100		
223	61.7	50.4	75.6
227	62.5	51.4	77.0



Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM







#35
Caprolactam
Concen: 10.65 ng
RT: 8.469 min Scan# 1085
Delta R.T. -0.035 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:113 Resp: 18525
Ion Ratio Lower Upper
113 100
55 233.4 254.9 294.9#
56 183.8 179.4 219.4

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#36
4-Chloro-3-methylphenol
Concen: 11.76 ng
RT: 8.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:107 Resp: 72684
Ion Ratio Lower Upper
107 100
144 22.5 16.3 24.5
142 64.2 50.1 75.1









#37
2-Methylnaphthalene
Concen: 11.58 ng
RT: 8.769 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:142 Resp: 132053
Ion Ratio Lower Upper
142 100
141 88.8 71.8 107.6




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#38
1-Methylnaphthalene
Concen: 11.99 ng
RT: 8.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01


Tgt Ion:142 Resp: 128514
Ion Ratio Lower Upper
142 100
141 88.2 72.8 109.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.810 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:164 Resp: 161935
Ion Ratio Lower Upper
164 100
162 109.0 80.7 121.1
160 48.1 37.3 55.9

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 12.84 ng
RT: 8.933 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:216 Resp: 61184
Ion Ratio Lower Upper
216 100
214 77.4 63.8 95.8
179 22.4 17.9 26.9
108 21.5 18.6 28.0







#41
Hexachlorocyclopentadiene
Concen: 5.67 ng
RT: 8.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion: 237 Resp: 10722
Ion Ratio Lower Upper
237 100
235 53.9 38.8 78.8
272 10.3 0.0 32.4




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#42
2,4,6-Tribromophenol
Concen: 26.20 ng
RT: 10.592 min Scan# 1446
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion: 330 Resp: 52820
Ion Ratio Lower Upper
330 100
332 96.7 77.1 115.7
141 33.3 27.9 41.9









#43
2,4,6-Trichlorophenol
Concen: 11.95 ng
RT: 9.051 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:196 Resp: 39227
Ion Ratio Lower Upper
196 100
198 103.3 78.2 117.2
200 31.5 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 11.43 ng
RT: 9.092 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:196 Resp: 40270
Ion Ratio Lower Upper
196 100
198 98.1 76.5 114.7
97 60.2 52.9 79.3
132 28.5 25.3 37.9



Instrument :
BNA_F
ClientSampled :
SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM







#45
2-Fluorobiphenyl
Concen: 25.89 ng
RT: 9.128 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:172 Resp: 287837
Ion Ratio Lower Upper
172 100
171 34.8 28.1 42.1
170 24.1 19.2 28.8

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#46
1,1'-Biphenyl
Concen: 11.97 ng
RT: 9.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01


Tgt Ion:154 Resp: 158748
Ion Ratio Lower Upper
154 100
153 40.7 20.7 60.7
76 10.9 0.0 31.7





#47
2-Chloronaphthalene
Concen: 12.40 ng
RT: 9.257 min Scan# 1219
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:162 Resp: 124031
Ion Ratio Lower Upper
162 100
127 42.1 36.2 54.2
164 32.9 25.4 38.2

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#48
2-Nitroaniline
Concen: 11.92 ng
RT: 9.351 min Scan# 1235
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion: 65 Resp: 52515
Ion Ratio Lower Upper
65 100
92 58.1 47.1 70.7
138 71.7 70.2 105.2





#49
Acenaphthylene
Concen: 12.38 ng
RT: 9.669 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:152 Resp: 199821
Ion Ratio Lower Upper
152 100
151 18.7 16.0 24.0
153 13.4 10.4 15.6

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#50
Dimethylphthalate
Concen: 11.73 ng
RT: 9.528 min Scan# 1265
Delta R.T. -0.012 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:163 Resp: 133779
Ion Ratio Lower Upper
163 100
194 4.5 3.1 4.7
164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 10.95 ng
RT: 9.592 min Scan# 1276
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Instrument :
BNA_F
ClientSampled :
SSTDICC010

Tgt Ion:165 Resp: 28887
Ion Ratio Lower Upper
165 100
63 100.1 67.7 101.5
89 85.1 58.8 88.2




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#52
Acenaphthene
Concen: 12.25 ng
RT: 9.839 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:154 Resp: 120500
Ion Ratio Lower Upper
154 100
153 113.8 88.1 132.1
152 52.7 43.4 65.0









#53
3-Nitroaniline
Concen: 11.21 ng
RT: 9.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:138 Resp: 35295
Ion Ratio Lower Upper
138 100
108 11.5 10.2 15.4
92 121.8 103.3 154.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM




#54
2,4-Dinitrophenol
Concen: 8.36 ng
RT: 9.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:184 Resp: 11709
Ion Ratio Lower Upper
184 100
63 85.4 73.5 110.3
154 82.5 52.5 78.7#







#55
Dibenzenofuran
Concen: 12.15 ng
RT: 10.010 min Scan# 1347
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:168 Resp: 180827
Ion Ratio Lower Upper
168 100
139 46.1 33.2 49.8
169 12.8 10.8 16.2




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#56
4-Nitrophenol
Concen: 9.22 ng
RT: 9.933 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:139 Resp: 21138
Ion Ratio Lower Upper
139 100
109 106.1 63.2 103.2#
65 145.3 110.0 150.0





#57
2,4-Dinitrotoluene
Concen: 11.08 ng
RT: 9.992 min Scan# 1344
Delta R.T. -0.012 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:165 Resp: 39824
Ion Ratio Lower Upper
165 100
63 76.1 43.5 65.3#
89 103.5 67.7 101.5#
182 2.0 2.1 3.1#

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#58
Fluorene
Concen: 12.28 ng
RT: 10.357 min Scan# 1406
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:166 Resp: 140434
Ion Ratio Lower Upper
166 100
165 95.3 73.5 110.3
167 14.5 10.7 16.1









#59
2,3,4,6-Tetrachlorophenol
Concen: 11.79 ng
RT: 10.133 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:232 Resp: 32870
Ion Ratio Lower Upper
232 100
131 55.6 43.2 64.8
130 2.7 2.2 3.2
166 33.9 27.3 40.9

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#60
Diethylphthalate
Concen: 11.66 ng
RT: 10.228 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01


Tgt Ion:149 Resp: 141535
Ion Ratio Lower Upper
149 100
177 17.7 15.4 23.2
150 12.1 9.9 14.9





#61
4-Chlorophenyl-phenylether
Concen: 12.58 ng
RT: 10.351 min Scan# 1405
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:204 Resp: 67054
Ion Ratio Lower Upper
204 100
206 31.9 25.7 38.5
141 61.9 55.5 83.3

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#62
4-Nitroaniline
Concen: 11.12 ng
RT: 10.369 min Scan# 1408
Delta R.T. -0.012 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:138 Resp: 34697
Ion Ratio Lower Upper
138 100
92 58.4 35.8 75.8
108 107.2 63.6 103.6#









#63
Azobenzene
Concen: 11.57 ng
RT: 10.510 min Scan# 1432
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:	77	Resp:	161498
Ion Ratio	100	Lower	Upper
182	22.0	1.7	41.7
105	18.3	1.7	41.7
51	39.1	15.9	55.9



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.292 min Scan# 1565
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:	188	Resp:	320159
Ion Ratio	100	Lower	Upper
94	7.4	6.9	10.3
80	8.8	8.2	12.2



Instrument :
BNA_F
ClientSampled :
SSTDIC010


Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#65
4,6-Dinitro-2-methylphenol
Concen: 9.81 ng
RT: 10.404 min Scan# 1414
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:198 Resp: 19456
Ion Ratio Lower Upper
198 100
51 85.4 64.8 104.8
105 61.9 42.6 82.6

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#66
n-Nitrosodiphenylamine
Concen: 11.76 ng
RT: 10.463 min Scan# 1424
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:169 Resp: 123138
Ion Ratio Lower Upper
169 100
168 62.4 53.3 79.9
167 36.2 29.9 44.9





#67
4-Bromophenyl-phenylether
Concen: 12.28 ng
RT: 10.833 min Scan# 1487
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010



Tgt Ion:248 Resp: 41398
Ion Ratio Lower Upper
248 100
250 96.5 77.4 116.2
141 96.4 63.7 95.5#




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM




#68
Hexachlorobenzene
Concen: 12.49 ng
RT: 10.904 min Scan# 1499
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:284 Resp: 47933
Ion Ratio Lower Upper
284 100
142 28.2 28.3 42.5#
249 30.0 23.5 35.3







#69
Atrazine
Concen: 11.14 ng
RT: 10.992 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:200 Resp: 35266
Ion Ratio Lower Upper
200 100
173 24.7 6.9 46.9
215 49.2 26.5 66.5




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#70
Pentachlorophenol
Concen: 8.70 ng
RT: 11.104 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:266 Resp: 17020
Ion Ratio Lower Upper
266 100
268 77.0 51.6 77.4
264 69.3 52.6 78.8









#71
Phenanthrene
Concen: 12.21 ng
RT: 11.316 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:178 Resp: 205048
Ion Ratio Lower Upper
178 100
176 18.1 15.6 23.4
179 14.7 12.5 18.7

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM




#72
Anthracene
Concen: 11.96 ng
RT: 11.369 min Scan# 1578
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:178 Resp: 199644
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.0
179 15.0 12.6 18.8









#73
Carbazole
Concen: 12.36 ng
RT: 11.522 min Scan# 1604
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:167 Resp: 208615
Ion Ratio Lower Upper
167 100
166 21.2 17.0 25.6
139 13.9 10.4 15.6



#74
Di-n-butylphthalate
Concen: 10.61 ng
RT: 11.857 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:149 Resp: 201052
Ion Ratio Lower Upper
149 100
150 8.7 7.8 11.8
104 5.4 4.6 6.8



Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#75
Fluoranthene
Concen: 11.97 ng
RT: 12.504 min Scan# 1771
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:202 Resp: 219235
Ion Ratio Lower Upper
202 100
101 7.1 0.0 30.9
203 17.0 0.0 37.5

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.933 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:240 Resp: 266479
Ion Ratio Lower Upper
240 100
120 6.6 7.2 10.8#
236 23.2 19.0 28.4





#77
Benzidine
Concen: 7.30 ng
RT: 12.627 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:184 Resp: 50373
Ion Ratio Lower Upper
184 100
185 14.6 12.1 18.1
183 10.1 7.8 11.8

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#78
Pyrene
Concen: 11.20 ng
RT: 12.733 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:202 Resp: 231614
Ion Ratio Lower Upper
202 100
200 19.1 16.1 24.1
203 17.7 14.8 22.2





#79
Terphenyl-d14
Concen: 23.63 ng
RT: 12.874 min Scan# 1834
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:244 Resp: 326947
Ion Ratio Lower Upper
244 100
212 7.4 5.7 8.5
122 7.7 9.7 14.5#

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#80
Butylbenzylphthalate
Concen: 9.36 ng
RT: 13.351 min Scan# 1915
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:149 Resp: 82706
Ion Ratio Lower Upper
149 100
91 76.0 63.6 95.4
206 19.1 14.5 21.7





#81
Benzo(a)anthracene
Concen: 11.89 ng
RT: 13.921 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:228 Resp: 192042
Ion Ratio Lower Upper
228 100
226 27.0 22.4 33.6
229 19.2 16.0 24.0

Manual Integrations
APPROVED



mohammad
5/11/2021 4:15:59 PM




#82
3,3'-Dichlorobenzidine
Concen: 11.07 ng
RT: 13.880 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01



Tgt Ion:252 Resp: 54906
Ion Ratio Lower Upper
252 100
254 65.4 51.1 76.7
126 8.6 7.7 11.5









#83
Chrysene
Concen: 11.56 ng
RT: 13.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:228 Resp: 187849
Ion Ratio Lower Upper
228 100
226 30.0 24.3 36.5
229 21.2 16.2 24.2

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 8.86 ng
RT: 13.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:149 Resp: 97291
Ion Ratio Lower Upper
149 100
167 25.0 21.8 32.6
279 4.1 3.3 4.9









#85
Di-n-octyl phthalate
Concen: 7.79 ng
RT: 14.527 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion:149 Resp: 136705
Ion Ratio Lower Upper
149 100
167 1.4 1.2 1.8
43 14.8 11.8 17.6



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.368 min Scan# 2258
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion:264 Resp: 185035
Ion Ratio Lower Upper
264 100
260 22.1 18.7 28.1
265 21.6 17.8 26.8




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED




mohammad
5/11/2021 4:15:59 PM




#87
Indeno(1,2,3-cd)pyrene
Concen: 10.99 ng
RT: 16.786 min Scan# 2499
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01





Tgt Ion: 276 Resp: 118481
Ion Ratio Lower Upper
276 100
138 14.7 16.2 24.2#
277 25.4 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 12.10 ng
RT: 14.951 min Scan# 2187
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01




Tgt Ion: 252 Resp: 150825
Ion Ratio Lower Upper
252 100
253 20.7 18.1 27.1
125 7.4 6.2 9.4



Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM







#89
Benzo(k)fluoranthene
Concen: 11.70 ng m
RT: 14.980 min Scan# 2192
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:252 Resp: 139265
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 6.4 7.0 10.6#

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
5/11/2021 4:15:59 PM



#90
Benzo(a)pyrene
Concen: 11.61 ng
RT: 15.310 min Scan# 2248
Delta R.T. 0.000 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion:252 Resp: 123795
Ion Ratio Lower Upper
252 100
253 20.7 18.4 27.6
125 7.2 8.0 12.0#









#91
Dibenzo(a,h)anthracene
Concen: 10.88 ng
RT: 16.798 min Scan# 2501
Delta R.T. -0.006 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 278 Resp: 99002
Ion Ratio Lower Upper
278 100
139 9.6 10.2 15.2#
279 24.6 19.4 29.0

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED

mohammad
5/11/2021 4:15:59 PM



#92
Benzo(g,h,i)perylene
Concen: 11.17 ng
RT: 17.209 min Scan# 2571
Delta R.T. -0.012 min
Lab File: BF123883.D
Acq: 10 May 2021 13:01

Tgt Ion: 276 Resp: 101814
Ion Ratio Lower Upper
276 100
277 21.9 18.2 27.2
138 14.9 13.8 20.6

