

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051022\
 Data File : BF128156.D
 Acq On : 10 May 2022 10:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 03:59:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 15:30:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	40.027	-0.1	103	0.00
3	Pyridine	40.000	40.641	-1.6	107	0.00
4	n-Nitrosodimethylamine	40.000	40.591	-1.5	104	0.00
5 S	2-Fluorophenol	80.000	77.948	2.6	104	0.00
6	Aniline	40.000	39.801	0.5	104	0.00
7 S	Phenol-d6	80.000	77.005	3.7	102	0.00
8	2-Chlorophenol	40.000	39.039	2.4	103	0.00
9	Benzaldehyde	40.000	41.149	-2.9	110	0.00
10 C	Phenol	40.000	39.064	2.3	101	0.00
11	bis(2-Chloroethyl)ether	40.000	39.394	1.5	104	0.00
12	1,3-Dichlorobenzene	40.000	39.284	1.8	103	0.00
13 C	1,4-Dichlorobenzene	40.000	39.437	1.4	103	0.00
14	1,2-Dichlorobenzene	40.000	38.523	3.7	102	0.00
15	Benzyl Alcohol	40.000	38.937	2.7	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.843	0.4	103	0.00
17	2-Methylphenol	40.000	38.420	3.9	101	0.00
18	Hexachloroethane	40.000	39.463	1.3	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.257	4.4	101	0.00
20	3+4-Methylphenols	40.000	38.427	3.9	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	39.258	1.9	102	0.00
23 S	Nitrobenzene-d5	80.000	80.150	-0.2	102	0.00
24	Nitrobenzene	40.000	40.518	-1.3	101	0.00
25	Isophorone	40.000	40.213	-0.5	101	0.00
26 C	2-Nitrophenol	40.000	40.752	-1.9	102	0.00
27	2,4-Dimethylphenol	40.000	40.242	-0.6	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.566	-1.4	103	0.00
29 C	2,4-Dichlorophenol	40.000	40.951	-2.4	103	0.00
30	1,2,4-Trichlorobenzene	40.000	40.562	-1.4	103	0.00
31	Naphthalene	40.000	39.759	0.6	101	0.00
32	Benzoic acid	40.000	42.205	-5.5	100	0.00
33	4-Chloroaniline	40.000	39.367	1.6	101	0.00
34 C	Hexachlorobutadiene	40.000	40.530	-1.3	103	0.00
35	Caprolactam	40.000	38.931	2.7	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.827	0.4	100	0.00
37	2-Methylnaphthalene	40.000	39.243	1.9	100	0.00
38	1-Methylnaphthalene	40.000	39.442	1.4	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.910	-2.3	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.111	-2.8	101	0.00
42 S	2,4,6-Tribromophenol	80.000	79.348	0.8	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.082	-2.7	100	0.00
44	2,4,5-Trichlorophenol	40.000	40.529	-1.3	99	0.00
45 S	2-Fluorobiphenyl	80.000	79.351	0.8	100	0.00
46	1,1'-Biphenyl	40.000	40.791	-2.0	101	0.00
47	2-Chloronaphthalene	40.000	40.276	-0.7	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051022\
 Data File : BF128156.D
 Acq On : 10 May 2022 10:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 03:59:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 15:30:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.609	-1.5	97	0.00
49	Acenaphthylene	40.000	40.156	-0.4	98	0.00
50	Dimethylphthalate	40.000	40.045	-0.1	99	0.00
51	2,6-Dinitrotoluene	40.000	40.540	-1.3	99	0.00
52 C	Acenaphthene	40.000	40.284	-0.7	99	0.00
53	3-Nitroaniline	40.000	39.726	0.7	97	0.00
54 P	2,4-Dinitrophenol	40.000	39.489	1.3	93	0.00
55	Dibenzofuran	40.000	39.522	1.2	99	0.00
56 P	4-Nitrophenol	40.000	41.398	-3.5	99	0.00
57	2,4-Dinitrotoluene	40.000	39.999	0.0	97	0.00
58	Fluorene	40.000	39.335	1.7	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.450	-1.1	98	0.00
60	Diethylphthalate	40.000	39.218	2.0	97	0.00
61	4-Chlorophenyl-phenylether	40.000	39.647	0.9	99	0.00
62	4-Nitroaniline	40.000	39.294	1.8	96	0.00
63	Azobenzene	40.000	39.981	0.0	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.161	-5.4	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.565	-1.4	98	0.00
67	4-Bromophenyl-phenylether	40.000	40.860	-2.1	99	0.00
68	Hexachlorobenzene	40.000	40.437	-1.1	98	0.00
69	Atrazine	40.000	40.663	-1.7	99	0.00
70 C	Pentachlorophenol	40.000	43.713	-9.3	95	0.00
71	Phenanthrene	40.000	39.933	0.2	98	0.00
72	Anthracene	40.000	39.642	0.9	96	0.00
73	Carbazole	40.000	39.084	2.3	94	0.00
74	Di-n-butylphthalate	40.000	40.160	-0.4	98	0.00
75 C	Fluoranthene	40.000	38.797	3.0	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	40.781	-2.0	94	0.00
78	Pyrene	40.000	41.299	-3.2	96	0.00
79 S	Terphenyl-d14	80.000	80.092	-0.1	94	0.00
80	Butylbenzylphthalate	40.000	41.670	-4.2	95	0.00
81	Benzo(a)anthracene	40.000	40.156	-0.4	93	0.00
82	3,3'-Dichlorobenzidine	40.000	40.210	-0.5	96	0.00
83	Chrysene	40.000	40.035	-0.1	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.600	-4.0	95	0.00
85 c	Di-n-octyl phthalate	40.000	41.340	-3.4	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.941	2.6	92	0.00
88	Benzo(b)fluoranthene	40.000	39.774	0.6	91	0.00
89	Benzo(k)fluoranthene	40.000	41.297	-3.2	93	0.00
90 C	Benzo(a)pyrene	40.000	40.278	-0.7	92	0.00
91	Dibenzo(a,h)anthracene	40.000	39.554	1.1	93	-0.01
92	Benzo(g,h,i)perylene	40.000	38.704	3.2	93	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051022\
Data File : BF128156.D
Acq On : 10 May 2022 10:23
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 11 03:59:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 15:30:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0