

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051024\
 Data File : BF137875.D
 Acq On : 10 May 2024 12:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: May 11 00:30:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.051	152	292576	20.000	ng	0.00
21) Naphthalene-d8	8.339	136	1117422	20.000	ng	0.00
39) Acenaphthene-d10	10.098	164	611685	20.000	ng	-0.01
64) Phenanthrene-d10	11.592	188	1008345	20.000	ng	0.00
76) Chrysene-d12	14.251	240	605529	20.000	ng	# 0.00
86) Perylene-d12	15.815	264	544671	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.675	112	1732477	99.142	ng	0.00
7) Phenol-d6	6.681	99	2230577	100.636	ng	0.00
23) Nitrobenzene-d5	7.622	82	1501142	78.049	ng	0.00
42) 2,4,6-Tribromophenol	10.898	330	695327	111.333	ng	0.00
45) 2-Fluorobiphenyl	9.416	172	2817919	74.739	ng	0.00
79) Terphenyl-d14	13.186	244	3092573	85.590	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.969	88	316711	40.248	ng	96
3) Pyridine	3.734	79	661033	34.709	ng	97
4) n-Nitrosodimethylamine	3.716	42	313383	36.542	ng	96
6) Aniline	6.722	93	841583	38.308	ng	100
8) 2-Chlorophenol	6.845	128	718987	37.755	ng	95
9) Benzaldehyde	6.610	77	433382	36.308	ng	99
10) Phenol	6.693	94	818798	36.052	ng	98
11) bis(2-Chloroethyl)ether	6.793	93	663567	37.314	ng	99
12) 1,3-Dichlorobenzene	6.998	146	832657	38.773	ng	97
13) 1,4-Dichlorobenzene	7.075	146	830589	38.546	ng	98
14) 1,2-Dichlorobenzene	7.228	146	761896	37.540	ng	99
15) Benzyl Alcohol	7.193	79	583178	38.046	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.322	45	860315	34.418	ng	97
17) 2-Methylphenol	7.293	107	600322	39.067	ng	98
18) Hexachloroethane	7.569	117	289853	39.790	ng	97
19) n-Nitroso-di-n-propyla...	7.469	70	466145	35.191	ng	96
20) 3+4-Methylphenols	7.451	107	748280	37.013	ng	90
22) Acetophenone	7.463	105	957069	38.083	ng	# 99
24) Nitrobenzene	7.640	77	768920	38.790	ng	98
25) Isophorone	7.875	82	1346514	38.829	ng	100
26) 2-Nitrophenol	7.951	139	398356	44.377	ng	94
27) 2,4-Dimethylphenol	7.981	122	504863	39.066	ng	99
28) bis(2-Chloroethoxy)met...	8.081	93	793576	37.796	ng	99
29) 2,4-Dichlorophenol	8.192	162	621843	39.810	ng	98
30) 1,2,4-Trichlorobenzene	8.275	180	686268	40.301	ng	99
31) Naphthalene	8.363	128	2124892	38.265	ng	99
32) Benzoic acid	8.110	122	464442	41.658	ng	95
33) 4-Chloroaniline	8.404	127	767669	37.280	ng	99
34) Hexachlorobutadiene	8.469	225	431297	41.254	ng	99
35) Caprolactam	8.792	113	162045	32.742	ng	95
36) 4-Chloro-3-methylphenol	8.881	107	633724	38.904	ng	97
37) 2-Methylnaphthalene	9.051	142	1370181	38.118	ng	99
38) 1-Methylnaphthalene	9.151	142	1328215	37.558	ng	100
40) 1,2,4,5-Tetrachloroben...	9.216	216	654599	40.405	ng	100
41) Hexachlorocyclopentadiene	9.198	237	271289	41.227	ng	99
43) 2,4,6-Trichlorophenol	9.328	196	436914	39.619	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051024\
 Data File : BF137875.D
 Acq On : 10 May 2024 12:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Quant Time: May 11 00:30:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.363	196	486968	39.991	ng	96
46) 1,1'-Biphenyl	9.516	154	1631065	38.854	ng	100
47) 2-Chloronaphthalene	9.545	162	1329062	39.215	ng	98
48) 2-Nitroaniline	9.639	65	361707	39.352	ng	94
49) Acenaphthylene	9.963	152	1857605	37.890	ng	100
50) Dimethylphthalate	9.816	163	1496799	38.762	ng	100
51) 2,6-Dinitrotoluene	9.881	165	359151	40.698	ng	96
52) Acenaphthene	10.139	154	1256451	38.664	ng	99
53) 3-Nitroaniline	10.057	138	352721	38.168	ng	97
54) 2,4-Dinitrophenol	10.151	184	186974	43.120	ng #	87
55) Dibenzofuran	10.310	168	1781086	37.663	ng	98
56) 4-Nitrophenol	10.198	139	280039	36.562	ng	98
57) 2,4-Dinitrotoluene	10.286	165	470843	40.840	ng	94
58) Fluorene	10.651	166	1405765	37.351	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.422	232	368680	37.293	ng	99
60) Diethylphthalate	10.516	149	1408230	38.398	ng	99
61) 4-Chlorophenyl-phenyle...	10.639	204	704641	37.997	ng	99
62) 4-Nitroaniline	10.675	138	350399	37.382	ng	96
63) Azobenzene	10.798	77	1270758	35.851	ng	98
65) 4,6-Dinitro-2-methylph...	10.692	198	262147	45.295	ng	98
66) n-Nitrosodiphenylamine	10.763	169	1193885	39.554	ng	99
67) 4-Bromophenyl-phenylether	11.133	248	449892	41.689	ng	99
68) Hexachlorobenzene	11.198	284	495606	42.025	ng	99
69) Atrazine	11.280	200	313449	34.058	ng	99
70) Pentachlorophenol	11.392	266	307339	37.960	ng	99
71) Phenanthrene	11.622	178	1859379	37.832	ng	100
72) Anthracene	11.675	178	1876452	38.372	ng	100
73) Carbazole	11.828	167	1816550	38.544	ng	100
74) Di-n-butylphthalate	12.145	149	2125518	41.344	ng	99
75) Fluoranthene	12.816	202	1996574	37.163	ng	99
77) Benzidine	12.927	184	338120	24.144	ng	99
78) Pyrene	13.045	202	2063182	42.973	ng	100
80) Butylbenzylphthalate	13.651	149	723483	47.494	ng	97
81) Benzo(a)anthracene	14.233	228	1524656	39.367	ng	100
82) 3,3'-Dichlorobenzidine	14.192	252	439175	37.879	ng	100
83) Chrysene	14.274	228	1420983	39.185	ng	99
84) Bis(2-ethylhexyl)phtha...	14.204	149	892124	46.470	ng	99
85) Di-n-octyl phthalate	14.833	149	1326923	51.358	ng	98
87) Indeno(1,2,3-cd)pyrene	17.457	276	1348297	43.541	ng	99
88) Benzo(b)fluoranthene	15.345	252	1308894	38.624	ng	99
89) Benzo(k)fluoranthene	15.374	252	1274049	38.897	ng	99
90) Benzo(a)pyrene	15.751	252	1075166	38.889	ng	99
91) Dibenzo(a,h)anthracene	17.480	278	1108044	43.960	ng	99
92) Benzo(g,h,i)perylene	17.962	276	1189718	45.025	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 11 00:30:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.051 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 292576
Ion Ratio Lower Upper
152 100
150 157.1 124.1 186.1
115 62.3 49.1 73.7



#2
1,4-Dioxane
Concen: 40.248 ng
RT: 2.969 min Scan# 150
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 88 Resp: 316711
Ion Ratio Lower Upper
88 100
58 69.5 58.8 88.2
43 24.9 21.4 32.0









#3
Pyridine
Concen: 34.709 ng
RT: 3.734 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 661033
Ion Ratio Lower Upper
79 100
52 67.5 55.5 83.3
51 31.8 26.9 40.3



#4
n-Nitrosodimethylamine
Concen: 36.542 ng
RT: 3.716 min Scan# 277
Delta R.T. 0.018 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 42 Resp: 313383
Ion Ratio Lower Upper
42 100
74 146.1 112.9 169.3
44 6.3 5.6 8.4









#5
2-Fluorophenol
Concen: 99.142 ng
RT: 5.675 min Scan# 610
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 1732477
Ion Ratio Lower Upper
112 100
64 56.5 47.3 70.9
63 29.3 23.7 35.5



#6
Aniline
Concen: 38.308 ng
RT: 6.722 min Scan# 788
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 93 Resp: 841583
Ion Ratio Lower Upper
93 100
66 39.6 31.8 47.6
65 18.9 15.4 23.2









#7
Phenol-d6
Concen: 100.636 ng
RT: 6.681 min Scan# 71
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 2230577




Ion	Ratio	Lower	Upper
99	100		
42	16.1	13.2	19.8
71	33.6	25.0	37.4




#8
2-Chlorophenol
Concen: 37.755 ng
RT: 6.845 min Scan# 809
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:128 Resp: 718987

Ion	Ratio	Lower	Upper
128	100		
130	33.3	12.5	52.5
64	39.2	23.9	63.9









#9
Benzaldehyde
Concen: 36.308 ng
RT: 6.610 min Scan# 770
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 433382
Ion Ratio Lower Upper
77 100
105 104.1 83.2 123.2
106 99.3 77.7 117.7



#10
Phenol
Concen: 36.052 ng
RT: 6.693 min Scan# 783
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 94 Resp: 818798
Ion Ratio Lower Upper
94 100
65 26.8 6.4 46.4
66 36.8 15.4 55.4









#11
bis(2-Chloroethyl)ether
Concen: 37.314 ng
RT: 6.793 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 663567
Ion Ratio Lower Upper
93 100
63 71.5 51.6 91.6
95 33.2 12.1 52.1



#12
1,3-Dichlorobenzene
Concen: 38.773 ng
RT: 6.998 min Scan# 835
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:146 Resp: 832657
Ion Ratio Lower Upper
146 100
148 65.0 51.2 76.8
75 28.5 25.1 37.7









#13
1,4-Dichlorobenzene
Concen: 38.546 ng
RT: 7.075 min Scan# 84
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 830589
Ion Ratio Lower Upper
146 100
148 64.6 50.7 76.1
111 40.1 33.8 50.8



#14
1,2-Dichlorobenzene
Concen: 37.540 ng
RT: 7.228 min Scan# 874
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:146 Resp: 761896
Ion Ratio Lower Upper
146 100
148 64.3 51.8 77.8
111 43.3 33.6 50.4









#15
Benzyl Alcohol
Concen: 38.046 ng
RT: 7.193 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 583178
Ion Ratio Lower Upper
79 100
108 79.8 64.2 96.2
77 61.9 48.5 72.7



#16
2,2'-oxybis(1-Chloropropane)
Concen: 34.418 ng
RT: 7.322 min Scan# 890
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 45 Resp: 860315
Ion Ratio Lower Upper
45 100
77 14.8 0.0 33.8
79 12.0 0.0 31.0









#17
2-Methylphenol
Concen: 39.067 ng
RT: 7.293 min Scan# 885
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:	107	Resp:	600322
Ion	Ratio	Lower	Upper
107	100		
108	110.6	90.5	135.7
77	43.7	34.5	51.7
79	44.1	35.2	52.8



#18
Hexachloroethane
Concen: 39.790 ng
RT: 7.569 min Scan# 932
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:	117	Resp:	289853
Ion	Ratio	Lower	Upper
117	100		
119	97.9	77.6	116.4
201	109.0	83.8	125.8









#19
n-Nitroso-di-n-propylamine
Concen: 35.191 ng
RT: 7.469 min Scan# 914
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 466145
Ion Ratio Lower Upper
70 100
42 53.4 44.6 66.8
101 10.6 7.8 11.6
130 24.9 17.5 26.3



#20
3+4-Methylphenols
Concen: 37.013 ng
RT: 7.451 min Scan# 912
Delta R.T. 0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 107 Resp: 748280
Ion Ratio Lower Upper
107 100
108 91.8 69.3 109.3
77 53.1 15.7 55.7
79 30.1 6.8 46.8









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.339 min Scan# 1063
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:136 Resp: 1117422
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.2 5.8 8.8
68 5.8 5.2 7.8



#22
Acetophenone
Concen: 38.083 ng
RT: 7.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:105 Resp: 957069
Ion Ratio Lower Upper
105 100
71 6.7 6.9 10.3#
51 25.0 20.2 30.2
120 23.3 18.6 27.8





#23

Nitrobenzene-d5

Concen: 78.049 ng

RT: 7.622 min Scan# 941

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 1501142

Ion Ratio Lower Upper

82 100

128 48.8 37.5 56.3

54 48.0 39.8 59.6



#24

Nitrobenzene

Concen: 38.790 ng

RT: 7.640 min Scan# 944

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion: 77 Resp: 768920


Ion Ratio Lower Upper

77 100

123 48.5 37.4 56.2

65 13.4 10.7 16.1









#25
Isophorone
Concen: 38.829 ng
RT: 7.875 min Scan# 991
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1346514
Ion Ratio Lower Upper
82 100
95 7.3 5.7 8.5
138 18.8 15.0 22.6



#26
2-Nitrophenol
Concen: 44.377 ng
RT: 7.951 min Scan# 997
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:139 Resp: 398356
Ion Ratio Lower Upper
139 100
109 28.2 20.2 30.2
65 45.0 33.4 50.0









#27
2,4-Dimethylphenol
Concen: 39.066 ng
RT: 7.981 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 504863
Ion Ratio Lower Upper
122 100
107 109.3 85.8 128.8
121 57.8 46.6 69.8



#28
bis(2-Chloroethoxy)methane
Concen: 37.796 ng
RT: 8.081 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 93 Resp: 793576
Ion Ratio Lower Upper
93 100
95 33.0 26.1 39.1
123 13.1 10.2 15.4





#29

2,4-Dichlorophenol

Concen: 39.810 ng

RT: 8.192 min Scan# 1038

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 621843

Ion Ratio Lower Upper

162 100

164 64.2 43.5 83.5

98 33.5 15.1 55.1



#30

1,2,4-Trichlorobenzene

Concen: 40.301 ng

RT: 8.275 min Scan# 1052

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:180 Resp: 686268

Ion Ratio Lower Upper

180 100

182 95.1 76.2 114.2

145 33.2 25.1 37.7









#31
Naphthalene
Concen: 38.265 ng
RT: 8.363 min Scan# 1067
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 2124892
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 13.5 10.6 16.0



#32
Benzoic acid
Concen: 41.658 ng
RT: 8.110 min Scan# 1024
Delta R.T. 0.024 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:122 Resp: 464442
Ion Ratio Lower Upper
122 100
105 119.4 105.3 145.3
77 88.8 73.2 113.2





#33

4-Chloroaniline

Concen: 37.280 ng

RT: 8.404 min Scan# 1074

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:127 Resp: 767669





Ion Ratio Lower Upper

127 100

129 31.8 25.7 38.5

65 28.9 23.4 35.2

92 20.4 15.6 23.4



#34

Hexachlorobutadiene

Concen: 41.254 ng

RT: 8.469 min Scan# 1085

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:225 Resp: 431297


Ion Ratio Lower Upper

225 100

223 63.5 51.1 76.7

227 62.9 51.4 77.0









#35
 Caprolactam
 Concen: 32.742 ng
 RT: 8.792 min Scan# 1138
 Delta R.T. 0.012 min
 Lab File: BF137875.D
 Acq: 10 May 2024 12:03




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 162045
 Ion Ratio Lower Upper
 113 100
 55 142.9 128.7 168.7
 56 107.7 93.9 133.9



#36
 4-Chloro-3-methylphenol
 Concen: 38.904 ng
 RT: 8.881 min Scan# 1155
 Delta R.T. 0.000 min
 Lab File: BF137875.D
 Acq: 10 May 2024 12:03

Tgt Ion:107 Resp: 633724
 Ion Ratio Lower Upper
 107 100
 144 26.0 21.4 32.0
 142 79.2 66.2 99.2





#37

2-Methylnaphthalene

Concen: 38.118 ng

RT: 9.051 min Scan# 1184

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:142 Resp: 1370181

Ion Ratio Lower Upper

142 100

141 88.0 70.8 106.2



#38

1-Methylnaphthalene

Concen: 37.558 ng

RT: 9.151 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BF137875.D




Acq: 10 May 2024 12:03


Tgt Ion:142 Resp: 1328215

Ion Ratio Lower Upper

142 100

141 91.7 73.1 109.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.098 min Scan# 11

Delta R.T. -0.012 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:164 Resp: 611685

Ion Ratio Lower Upper

164 100

162 98.4 79.8 119.6

160 41.4 34.4 51.6



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.405 ng

RT: 9.216 min Scan# 1212

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Tgt Ion:216 Resp: 654599




Ion Ratio Lower Upper


216 100

214 78.7 63.3 94.9

179 20.9 16.8 25.2

108 18.0 14.3 21.5









#41
Hexachlorocyclopentadiene
Concen: 41.227 ng
RT: 9.198 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:237 Resp: 271289
Ion Ratio Lower Upper
237 100
235 62.2 43.0 83.0
272 12.6 0.0 32.8



#42
2,4,6-Tribromophenol
Concen: 111.333 ng
RT: 10.898 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:330 Resp: 695327
Ion Ratio Lower Upper
330 100
332 95.7 77.3 115.9
141 26.7 23.8 35.6





#43

2,4,6-Trichlorophenol

Concen: 39.619 ng

RT: 9.328 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:196 Resp: 436914

Ion Ratio Lower Upper

196 100

198 96.6 76.2 114.4

200 31.3 24.2 36.2



#44

2,4,5-Trichlorophenol

Concen: 39.991 ng

RT: 9.363 min Scan# 1237

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Tgt Ion:196 Resp: 486968




Ion Ratio Lower Upper


196 100

198 96.6 76.6 114.8

97 49.1 34.5 51.7

132 28.7 20.3 30.5





#45

2-Fluorobiphenyl

Concen: 74.739 ng

RT: 9.416 min Scan# 1246

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:172 Resp: 2817919

Ion Ratio Lower Upper

172 100

171 37.0 29.6 44.4

170 24.6 19.4 29.2



#46

1,1'-Biphenyl

Concen: 38.854 ng

RT: 9.516 min Scan# 1263

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:154 Resp: 1631065


Ion Ratio Lower Upper

154 100

153 41.4 21.4 61.4

76 13.2 0.0 33.5





#47

2-Chloronaphthalene

Concen: 39.215 ng

RT: 9.545 min Scan# 1268

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 1329062

Ion Ratio Lower Upper

162 100

127 42.4 32.6 49.0

164 32.4 26.5 39.7



#48

2-Nitroaniline

Concen: 39.352 ng

RT: 9.639 min Scan# 1284

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion: 65 Resp: 361707


Ion Ratio Lower Upper

65 100

92 72.6 58.6 88.0

138 122.2 90.3 135.5





#49

Acenaphthylene

Concen: 37.890 ng

RT: 9.963 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:152 Resp: 1857605

Ion Ratio Lower Upper

152 100

151 20.6 16.2 24.4

153 13.5 10.8 16.2



#50

Dimethylphthalate

Concen: 38.762 ng

RT: 9.816 min Scan# 1314

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:163 Resp: 1496799


Ion Ratio Lower Upper

163 100

194 4.3 3.3 4.9

164 10.4 8.3 12.5









#51
2,6-Dinitrotoluene
Concen: 40.698 ng
RT: 9.881 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 359151
Ion Ratio Lower Upper
165 100
63 48.9 41.9 62.9
89 51.4 43.2 64.8



#52
Acenaphthene
Concen: 38.664 ng
RT: 10.139 min Scan# 1369
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:154 Resp: 1256451
Ion Ratio Lower Upper
154 100
153 111.5 88.7 133.1
152 52.9 42.8 64.2









#53
3-Nitroaniline
Concen: 38.168 ng
RT: 10.057 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 352721
Ion Ratio Lower Upper
138 100
108 10.3 7.8 11.6
92 102.0 84.6 126.8



#54
2,4-Dinitrophenol
Concen: 43.120 ng
RT: 10.151 min Scan# 1371
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:184 Resp: 186974
Ion Ratio Lower Upper
184 100
63 60.5 40.3 60.5#
154 65.4 45.2 67.8









#55
Dibenzofuran
Concen: 37.663 ng
RT: 10.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1781086
Ion Ratio Lower Upper
168 100
139 37.0 30.2 45.4
169 13.3 11.4 17.0



#56
4-Nitrophenol
Concen: 36.562 ng
RT: 10.198 min Scan# 1379
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:139 Resp: 280039
Ion Ratio Lower Upper
139 100
109 66.1 44.0 84.0
65 86.3 64.3 104.3









#57
2,4-Dinitrotoluene
Concen: 40.840 ng
RT: 10.286 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 470843
Ion Ratio Lower Upper
165 100
63 36.0 31.3 46.9
89 62.0 54.2 81.4
182 2.8 2.2 3.2



#58
Fluorene
Concen: 37.351 ng
RT: 10.651 min Scan# 1456
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:166 Resp: 1405765
Ion Ratio Lower Upper
166 100
165 96.4 78.2 117.2
167 13.8 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 37.293 ng
RT: 10.422 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 368680
Ion Ratio Lower Upper
232 100
131 42.3 33.4 50.0
130 2.2 1.7 2.5
166 29.4 22.8 34.2



#60
Diethylphthalate
Concen: 38.398 ng
RT: 10.516 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:149 Resp: 1408230
Ion Ratio Lower Upper
149 100
177 23.4 18.4 27.6
150 12.9 10.5 15.7









#61
4-Chlorophenyl-phenylether
Concen: 37.997 ng
RT: 10.639 min Scan# 1455
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 704641
Ion Ratio Lower Upper
204 100
206 32.4 26.4 39.6
141 59.9 47.0 70.6



#62
4-Nitroaniline
Concen: 37.382 ng
RT: 10.675 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:138 Resp: 350399
Ion Ratio Lower Upper
138 100
92 48.1 26.8 66.8
108 68.7 44.7 84.7









#63
Azobenzene
Concen: 35.851 ng
RT: 10.798 min Scan# 1482
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 1270758
Ion Ratio Lower Upper
77 100
182 26.5 5.1 45.1
105 22.9 2.2 42.2
51 28.0 8.7 48.7



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.592 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion: 188 Resp: 1008345
Ion Ratio Lower Upper
188 100
94 10.4 8.8 13.2
80 11.0 9.6 14.4









#65
4,6-Dinitro-2-methylphenol
Concen: 45.295 ng
RT: 10.692 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 262147
Ion Ratio Lower Upper
198 100
51 43.2 25.1 65.1
105 43.4 24.1 64.1



#66
n-Nitrosodiphenylamine
Concen: 39.554 ng
RT: 10.763 min Scan# 1475
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:169 Resp: 1193885
Ion Ratio Lower Upper
169 100
168 67.2 53.3 79.9
167 35.7 29.2 43.8









#67
4-Bromophenyl-phenylether
Concen: 41.689 ng
RT: 11.133 min Scan# 1538
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 449892
Ion Ratio Lower Upper
248 100
250 97.1 79.1 118.7
141 70.0 55.4 83.2



#68
Hexachlorobenzene
Concen: 42.025 ng
RT: 11.198 min Scan# 1549
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:284 Resp: 495606
Ion Ratio Lower Upper
284 100
142 33.4 27.3 40.9
249 32.4 25.6 38.4





#69

Atrazine

Concen: 34.058 ng

RT: 11.280 min Scan# 1563

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:200 Resp: 313449

Ion Ratio Lower Upper

200 100

173 27.3 5.9 45.9

215 48.6 28.8 68.8



#70

Pentachlorophenol

Concen: 37.960 ng

RT: 11.392 min Scan# 1582

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:266 Resp: 307339


Ion Ratio Lower Upper

266 100

268 64.1 51.3 76.9

264 62.5 51.0 76.4









#71
Phenanthrene
Concen: 37.832 ng
RT: 11.622 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1859379
Ion Ratio Lower Upper
178 100
176 19.7 15.8 23.6
179 15.7 12.5 18.7



#72
Anthracene
Concen: 38.372 ng
RT: 11.675 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:178 Resp: 1876452
Ion Ratio Lower Upper
178 100
176 19.3 15.2 22.8
179 15.8 12.6 18.8





#73

Carbazole

Concen: 38.544 ng

RT: 11.828 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:167 Resp: 1816550

Ion Ratio Lower Upper

167 100

166 22.0 17.6 26.4

139 12.9 10.5 15.7



#74

Di-n-butylphthalate

Concen: 41.344 ng

RT: 12.145 min Scan# 1710

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:149 Resp: 2125518


Ion Ratio Lower Upper

149 100

150 9.1 7.7 11.5

104 4.7 3.8 5.8





#75

Fluoranthene

Concen: 37.163 ng

RT: 12.816 min Scan# 1824

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:202 Resp: 1996574

Ion Ratio Lower Upper

202 100

101 10.9 0.0 31.8

203 17.7 0.0 37.8



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.251 min Scan# 2068

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:240 Resp: 605529


Ion Ratio Lower Upper

240 100

120 9.6 10.4 15.6#

236 25.7 20.4 30.6





#77

Benzidine

Concen: 24.144 ng

RT: 12.927 min Scan# 1843

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:184 Resp: 338120

Ion Ratio Lower Upper

184 100

185 15.2 12.3 18.5

183 12.3 9.2 13.8



#78

Pyrene

Concen: 42.973 ng

RT: 13.045 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:202 Resp: 2063182


Ion Ratio Lower Upper

202 100

200 21.4 17.2 25.8

203 17.7 14.4 21.6





#79

Terphenyl-d14

Concen: 85.590 ng

RT: 13.186 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:244 Resp: 3092573

Ion Ratio Lower Upper

244 100

212 6.7 5.5 8.3

122 9.1 9.4 14.2#



#80

Butylbenzylphthalate

Concen: 47.494 ng

RT: 13.651 min Scan# 1966

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:149 Resp: 723483


Ion Ratio Lower Upper

149 100

91 68.0 56.8 85.2

206 22.6 17.8 26.8









#81
 Benzo(a)anthracene
 Concen: 39.367 ng
 RT: 14.233 min Scan# 2065
 Delta R.T. -0.006 min
 Lab File: BF137875.D
 Acq: 10 May 2024 12:03

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:228 Resp: 1524656




Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.2	33.2
229	20.2	16.3	24.5




#82
 3,3'-Dichlorobenzidine
 Concen: 37.879 ng
 RT: 14.192 min Scan# 2058
 Delta R.T. -0.006 min
 Lab File: BF137875.D
 Acq: 10 May 2024 12:03

Tgt Ion:252 Resp: 439175

Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.7	77.5
126	12.1	9.7	14.5





#83

Chrysene

Concen: 39.185 ng

RT: 14.274 min Scan# 2073

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:228 Resp: 1420983

Ion Ratio Lower Upper

228 100

226 30.5 24.8 37.2

229 19.7 16.0 24.0



#84

Bis(2-ethylhexyl)phthalate

Concen: 46.470 ng

RT: 14.204 min Scan# 2060

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:149 Resp: 892124


Ion Ratio Lower Upper

149 100

167 28.7 22.4 33.6

279 5.0 3.6 5.4





#85

Di-n-octyl phthalate

Concen: 51.358 ng

RT: 14.833 min Scan# 2168

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:149 Resp: 1326923

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9

43 8.6 7.5 11.3



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.815 min Scan# 2334

Delta R.T. -0.006 min

Lab File: BF137875.D

Acq: 10 May 2024 12:03




Tgt Ion:264 Resp: 544671


Ion Ratio Lower Upper

264 100

260 22.9 18.9 28.3

265 20.9 17.4 26.0









#87
Indeno(1,2,3-cd)pyrene
Concen: 43.541 ng
RT: 17.457 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 1348297
Ion Ratio Lower Upper
276 100
138 25.7 21.4 32.0
277 25.5 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 38.624 ng
RT: 15.345 min Scan# 2254
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:252 Resp: 1308894
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.4 7.9 11.9









#89
Benzo(k)fluoranthene
Concen: 38.897 ng
RT: 15.374 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 1274049
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 11.0 9.0 13.4



#90
Benzo(a)pyrene
Concen: 38.889 ng
RT: 15.751 min Scan# 2323
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:252 Resp: 1075166
Ion Ratio Lower Upper
252 100
253 21.7 17.1 25.7
125 10.2 9.0 13.6









#91
Dibenzo(a,h)anthracene
Concen: 43.960 ng
RT: 17.480 min Scan# 2618
Delta R.T. -0.006 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 1108044
Ion Ratio Lower Upper
278 100
139 16.4 13.0 19.6
279 23.9 18.7 28.1



#92
Benzo(g,h,i)perylene
Concen: 45.025 ng
RT: 17.962 min Scan# 2699
Delta R.T. 0.000 min
Lab File: BF137875.D
Acq: 10 May 2024 12:03

Tgt Ion:276 Resp: 1189718
Ion Ratio Lower Upper
276 100
277 24.3 18.0 27.0
138 22.6 18.1 27.1

