


Data Path : Z:\HPCHEM1\BNA\_F\Data\BF051115\  
 Data File : BF079109.D  
 Acq On : 11 May 2015 11:56  
 Operator : TP/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF043015.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Apr 30 19:07:47 2015



AutoFind: Scans 1056, 1057, 1058; Background Corrected with Scan 1053

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 39.9      | 7805    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 34.5      | 6762    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 48.6      | 9514    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 19577   | PASS             |
| 199         | 198          | 5            | 9            | 6.3       | 1237    | PASS             |
| 275         | 198          | 10           | 60           | 24.3      | 4760    | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 648     | PASS             |
| 441         | 442          | 0.01         | 24           | 16.3      | 2588    | PASS             |
| 442         | 198          | 50           | 100          | 81.3      | 15911   | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 3013    | PASS             |