



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051115\
Data File : BF079133.D
Acq On : 12 May 2015 00:56
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 06:45:35 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 05:22:05 2015
Response via : Initial Calibration



TIC: BF079133.D

(69) Pentachlorophenol (C)



12.662min (+0.000) 0.00ng

response 24192

Ion	Exp%	Act%
265.70	100	100
268.00	68.80	69.05
264.00	60.20	60.73
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051115\
Data File : BF079133.D
Acq On : 12 May 2015 00:56
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 06:45:35 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 05:22:05 2015
Response via : Initial Calibration



TIC: BF079133.D

(76) Benzidine

14.582min (+0.000) 0.00ng

response 257481

Ion	Exp%	Act%
184.00	100	100
185.00	15.40	13.83
183.00	10.50	11.02
0.00	0.00	0.00