


Data Path : Z:\HPCHEM1\BNA F\Data\BF051115\
 Data File : BF079109.D
 Acq On : 11 May 2015 11:56
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 30 19:07:47 2015



AutoFind: Scans 1056, 1057, 1058; Background Corrected with Scan 1053

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	7805	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.5	6762	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.6	9514	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	19577	PASS
199	198	5	9	6.3	1237	PASS
275	198	10	60	24.3	4760	PASS
365	198	1	100	3.3	648	PASS
441	442	0.01	24	16.3	2588	PASS
442	198	50	100	81.3	15911	PASS
443	442	15	24	18.9	3013	PASS