


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051116\
 Data File : BF086895.D
 Acq On : 11 May 2016 17:45
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050916.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 09 16:04:56 2016



AutoFind: Scans 859, 860, 861; Background Corrected with Scan 856

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	69.9	33505	PASS
68	69	0.00	2	1.6	372	PASS
69	198	0.00	100	49.0	23483	PASS
70	69	0.00	2	0.6	130	PASS
127	198	10	80	55.3	26480	PASS
197	198	0.00	2	0.7	327	PASS
198	198	100	100	100.0	47907	PASS
199	198	5	9	6.3	2996	PASS
275	198	10	60	24.7	11823	PASS
365	198	1	100	3.9	1878	PASS
441	442	0.01	24	15.8	6733	PASS
442	198	50	100	88.9	42598	PASS
443	442	15	24	18.5	7875	PASS