

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
 Data File : BF095097.D
 Acq On : 12 May 2017 00:25
 Operator : SJ/MA
 Sample : I3058-04 20X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KY032MW006-170508

Manual Integrations
 APPROVED

mohammad
 5/12/2017 3:36:23 PM

Quant Time: May 12 13:14:10 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 17:24:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	85944	20.00	ng	-0.03
21) Naphthalene-d8	9.72	136	341572	20.00	ng	-0.04
38) Acenaphthene-d10	12.55	164	161628	20.00	ng	-0.04
63) Phenanthrene-d10	14.95	188	276096	20.00	ng	-0.03
75) Chrysene-d12	18.65	240	182071	20.00	ng	0.00
86) Perylene-d12	20.31	264	169692	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.74	112	1199	0.23	ng	0.06
7) Phenol-d6	7.44	99	1851	0.30	ng	0.16
23) Nitrobenzene-d5	8.61	82	361	0.07	ng	-0.03
41) 2,4,6-Tribromophenol	13.91	330	111	0.08	ng	0.02
44) 2-Fluorobiphenyl	11.51	172	5041	0.45	ng	-0.02
78) Terphenyl-d14	17.30	244	2830	0.34	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	2.03	88	30311	11.99	ng	# 90
14) 1,2-Dichlorobenzene	7.97	146	31347	4.98	ng	98
20) 3+4-Methylphenols	8.50	107	36801m	5.64	ng	
27) 2,4-Dimethylphenol	9.35	122	12337	2.49	ng	# 91
83) Bis(2-ethylhexyl)phthalate	18.70	149	18385	2.04	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
Data File : BF095097.D
Acq On : 12 May 2017 00:25
Operator : SJ/MA
Sample : I3058-04 20X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
KY032MW006-170508

Manual Integrations
APPROVED

mohammad
5/12/2017 3:36:23 PM

Quant Time: May 12 13:14:10 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 17:24:51 2017
Response via : Initial Calibration

