

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
 Data File : BF095098.D
 Acq On : 12 May 2017 00:57
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 12 04:40:54 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 17:24:51 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	-0.03
2	1,4-Dioxane	40.000	38.905	2.7	101	-0.08
3	Pyridine	40.000	38.661	3.3	98	-0.07
4	n-Nitrosodimethylamine	40.000	35.522	11.2	91	-0.07
5 S	2-Fluorophenol	80.000	78.503	1.9	98	-0.04
6	Aniline	40.000	41.288	-3.2	97	-0.03
7 S	Phenol-d6	80.000	83.241	-4.1	103	-0.03
8	2-Chlorophenol	40.000	40.032	-0.1	100	-0.02
9	Benzaldehyde	40.000	40.382	-1.0	98	-0.02
10 C	Phenol	40.000	44.088	-10.2	109	-0.03
11	bis(2-Chloroethyl)ether	40.000	40.824	-2.1	101	-0.03
12	1,3-Dichlorobenzene	40.000	39.795	0.5	106	-0.04
13 C	1,4-Dichlorobenzene	40.000	40.597	-1.5	100	-0.04
14	1,2-Dichlorobenzene	40.000	40.515	-1.3	101	-0.04
15	Benzyl Alcohol	40.000	45.208	-13.0	107	-0.02
16	2,2'-oxybis(1-Chloropropane)	40.000	40.454	-1.1	103	-0.02
17	2-Methylphenol	40.000	44.004	-10.0	113	-0.02
18	Hexachloroethane	40.000	39.923	0.2	98	-0.04
19 P	n-Nitroso-di-n-propylamine	40.000	40.557	-1.4	103	-0.02
20	3+4-Methylphenols	40.000	40.771	-1.9	101	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	104	-0.03
22	Acetophenone	40.000	38.746	3.1	102	-0.02
23 S	Nitrobenzene-d5	80.000	79.412	0.7	103	-0.02
24	Nitrobenzene	40.000	39.430	1.4	105	-0.02
25	Isophorone	40.000	38.582	3.5	101	-0.02
26 C	2-Nitrophenol	40.000	38.863	2.8	99	-0.02
27	2,4-Dimethylphenol	40.000	38.733	3.2	105	-0.02
28	bis(2-Chloroethoxy)methane	40.000	39.960	0.1	105	-0.02
29 C	2,4-Dichlorophenol	40.000	39.932	0.2	108	-0.01
30	1,2,4-Trichlorobenzene	40.000	39.718	0.7	106	-0.02
31	Naphthalene	40.000	39.358	1.6	105	-0.03
32	Benzoic acid	40.000	13.281	66.8#	27	-0.05
33	4-Chloroaniline	40.000	41.231	-3.1	108	-0.02
34 C	Hexachlorobutadiene	40.000	38.089	4.8	102	-0.04
35	Caprolactam	40.000	37.698	5.8	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.906	-2.3	107	-0.01
37	2-Methylnaphthalene	40.000	39.839	0.4	106	-0.03
38 I	Acenaphthene-d10	20.000	20.000	0.0	107	-0.03
39	1,2,4,5-Tetrachlorobenzene	40.000	40.453	-1.1	103	-0.03
40 P	Hexachlorocyclopentadiene	40.000	34.756	13.1	89	-0.04
41 S	2,4,6-Tribromophenol	80.000	78.025	2.5	98	-0.03
42 C	2,4,6-Trichlorophenol	40.000	40.309	-0.8	103	-0.02
43	2,4,5-Trichlorophenol	40.000	36.838	7.9	94	0.00
44 S	2-Fluorobiphenyl	80.000	77.058	3.7	101	-0.03

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
 Data File : BF095098.D
 Acq On : 12 May 2017 00:57
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 12 04:40:54 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 17:24:51 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.909	-2.3	104	-0.03
46	2-Chloronaphthalene	40.000	40.631	-1.6	106	-0.02
47	2-Nitroaniline	40.000	42.542	-6.4	112	-0.02
48	Acenaphthylene	40.000	38.176	4.6	100	-0.03
49	Dimethylphthalate	40.000	41.183	-3.0	107	-0.03
50	2,6-Dinitrotoluene	40.000	39.392	1.5	101	-0.02
51 C	Acenaphthene	40.000	39.431	1.4	103	-0.04
52	3-Nitroaniline	40.000	43.271	-8.2	110	-0.01
53 P	2,4-Dinitrophenol	40.000	7.079	82.3#	3	0.05
54	Dibenzofuran	40.000	41.619	-4.0	110	-0.03
55 P	4-Nitrophenol	40.000	31.799	20.5	78	0.02
56	2,4-Dinitrotoluene	40.000	43.108	-7.8	109	-0.01
57	Fluorene	40.000	40.190	-0.5	108	-0.03
58	2,3,4,6-Tetrachlorophenol	40.000	38.784	3.0	102	-0.02
59	Diethylphthalate	40.000	40.895	-2.2	110	-0.03
60	4-Chlorophenyl-phenylether	40.000	41.616	-4.0	108	-0.04
61	4-Nitroaniline	40.000	43.483	-8.7	115	-0.01
62	Azobenzene	40.000	39.533	1.2	101	-0.04
63 I	Phenanthrene-d10	20.000	20.000	0.0	106	-0.03
64	4,6-Dinitro-2-methylphenol	40.000	8.014	80.0#	21	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.478	-1.2	108	-0.03
66	4-Bromophenyl-phenylether	40.000	39.851	0.4	103	-0.04
67	Hexachlorobenzene	40.000	40.001	-0.0	106	-0.03
68	Atrazine	40.000	41.004	-2.5	114	-0.02
69 C	Pentachlorophenol	40.000	24.146	39.6#	64	-0.01
70	Phenanthrene	40.000	39.390	1.5	107	-0.03
71	Anthracene	40.000	40.214	-0.5	108	-0.03
72	Carbazole	40.000	40.114	-0.3	109	-0.02
73	Di-n-butylphthalate	40.000	43.497	-8.7	114	-0.03
74 C	Fluoranthene	40.000	38.363	4.1	101	-0.03
75 I	Chrysene-d12	20.000	20.000	0.0	89	-0.02
76	Benzidine	40.000	21.341	46.6#	39	-0.01
77	Pyrene	40.000	46.271	-15.7	101	-0.02
78 S	Terphenyl-d14	80.000	92.139	-15.2	103	-0.03
79	Butylbenzylphthalate	40.000	46.726	-16.8	101	-0.03
80	Benzo(a)anthracene	40.000	39.178	2.1	87	-0.02
81	3,3'-Dichlorobenzidine	40.000	33.645	15.9	72	-0.03
82	Chrysene	40.000	39.817	0.5	87	-0.03
83	Bis(2-ethylhexyl)phthalate	40.000	46.920	-17.3	101	-0.04
84 c	Di-n-octyl phthalate	40.000	43.657	-9.1	94	-0.04
85	Indeno(1,2,3-cd)pyrene	40.000	47.748	-19.4	108	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	95	-0.02
87	Benzo(b)fluoranthene	40.000	37.765	5.6	83	-0.02

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
Data File : BF095098.D
Acq On : 12 May 2017 00:57
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 12 04:40:54 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 17:24:51 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.147	-0.4	92	-0.02
89 C	Benzo(a)pyrene	40.000	38.363	4.1	87	-0.02
90	Dibenzo(a,h)anthracene	40.000	45.708	-14.3	106	-0.03
91	Benzo(a,h,i)perylene	40.000	49.975	-24.9	119	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1