


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\
 Data File : BF095077.D
 Acq On : 11 May 2017 12:55
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2305, 2306, 2307; Background Corrected with Scan 2299

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	22322	PASS
68	69	0.00	2	1.5	353	PASS
69	198	0.00	100	35.9	23665	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.6	32088	PASS
197	198	0.00	2	0.5	310	PASS
198	198	100	100	100.0	65965	PASS
199	198	5	9	6.9	4523	PASS
275	198	10	60	22.8	15059	PASS
365	198	1	100	2.3	1540	PASS
441	442	0.01	24	15.8	8862	PASS
442	198	50	100	85.3	56258	PASS
443	442	15	24	20.4	11464	PASS