


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051117\  
 Data File : BF095099.D  
 Acq On : 12 May 2017 2:01  
 Operator : SJ/MA  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF050217.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2304, 2305, 2306; Background Corrected with Scan 2299

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 35.9      | 20669   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 292     | PASS             |
| 69          | 198          | 0.00         | 100          | 37.6      | 21632   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 50.8      | 29224   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 57530   | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 3866    | PASS             |
| 275         | 198          | 10           | 60           | 21.9      | 12588   | PASS             |
| 365         | 198          | 1            | 100          | 2.4       | 1361    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.1      | 7132    | PASS             |
| 442         | 198          | 50           | 100          | 77.1      | 44357   | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 8684    | PASS             |