

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051215\
 Data File : BF079187.D
 Acq On : 13 May 2015 7:14
 Operator : TP/IZ
 Sample : G2194-11
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TU021-SB-19-2.5

Quant Time: May 13 13:14:26 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 07:34:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.28	152	49664	20.00	ng	0.00
21) Naphthalene-d8	8.86	136	211494	20.00	ng	0.00
38) Acenaphthene-d10	11.03	164	101343	20.00	ng	-0.01
63) Phenanthrene-d10	12.88	188	201996	20.00	ng	0.00
75) Chrysene-d12	16.17	240	208600	20.00	ng	0.00
86) Perylene-d12	17.94	264	223495	20.00	ng	0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	330635	114.60	ng	0.00
7) Phenol-d6	6.84	99	424534	111.32	ng	0.00
23) Nitrobenzene-d5	7.98	82	253933	69.00	ng	0.00
41) 2,4,6-Tribromophenol	12.02	330	119896	109.36	ng	0.00
44) 2-Fluorobiphenyl	10.20	172	482126	66.28	ng	0.00
78) Terphenyl-d14	14.88	244	535751	61.49	ng	0.00
Target Compounds						
32) Benzoic acid	8.52	122	9694	10.76	ng	Qvalue 92
48) Acenaphthylene	10.86	152	22356	2.17	ng	100
49) Dimethylphthalate	10.71	163	58599	7.88	ng	99
74) Fluoranthene	14.39	202	52192	4.43	ng	96
77) Pyrene	14.66	202	66483	5.53	ng	97
80) Benzo(a)anthracene	16.16	228	34764	3.04	ng	# 77
82) Chrysene	16.21	228	32235	2.93	ng	# 96
87) Benzo(b)fluoranthene	17.49	252	53080m	4.34	ng	
89) Benzo(a)pyrene	17.87	252	35187	3.12	ng	# 90
91) Benzo(g,h,i)perylene	19.83	276	29827	2.49	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051215\
Data File : BF079187.D
Acq On : 13 May 2015 7:14
Operator : TP/IZ
Sample : G2194-11
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TU021-SB-19-2.5

Quant Time: May 13 13:14:26 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 13 07:34:34 2015
Response via : Initial Calibration

