

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051216\
 Data File : BF086945.D
 Acq On : 12 May 2016 22:20
 Operator : UM/SJ
 Sample : H2994-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PM-STA-1343-28.39(0-4)

Manual Integrations
 APPROVED

sohil
 5/13/2016 8:05:07 PM

Quant Time: May 13 02:00:55 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 00:58:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.62	152	137801	20.00	ng	-0.01
21) Naphthalene-d8	7.91	136	560759	20.00	ng	-0.01
38) Acenaphthene-d10	9.66	164	254810	20.00	ng	-0.01
63) Phenanthrene-d10	11.13	188	404682	20.00	ng	-0.01
75) Chrysene-d12	13.76	240	276838	20.00	ng	-0.01
86) Perylene-d12	15.12	264	292001	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	849767	104.43	ng	0.00
7) Phenol-d6	6.28	99	1077224	99.06	ng	-0.01
23) Nitrobenzene-d5	7.20	82	790100	70.75	ng	-0.01
41) 2,4,6-Tribromophenol	10.44	330	224138	83.09	ng	-0.01
44) 2-Fluorobiphenyl	8.99	172	1246217	82.20	ng	0.00
78) Terphenyl-d14	12.72	244	839210	64.32	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.39	163	82268	4.23	ng	97
74) Fluoranthene	12.34	202	59895	2.70	ng	92
77) Pyrene	12.56	202	53722	2.53	ng	97
80) Benzo(a)anthracene	13.75	228	51393	3.31	ng	# 93
85) Indeno(1,2,3-cd)pyrene	16.37	276	27386	2.08	ng	# 86
87) Benzo(b)fluoranthene	14.75	252	64835m	3.42	ng	#
89) Benzo(a)pyrene	15.06	252	36547	2.23	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051216\
Data File : BF086945.D
Acq On : 12 May 2016 22:20
Operator : UM/SJ
Sample : H2994-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PM-STA-1343-28.39(0-4)

Manual Integrations
APPROVED

sohil
5/13/2016 8:05:07 PM

Quant Time: May 13 02:00:55 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 13 00:58:18 2016
Response via : Initial Calibration

