


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051217\
 Data File : BF095108.D
 Acq On : 12 May 2017 13:03
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2304, 2305, 2306; Background Corrected with Scan 2298

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	13734	PASS
68	69	0.00	2	0.8	112	PASS
69	198	0.00	100	34.6	14253	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.6	19585	PASS
197	198	0.00	2	0.5	218	PASS
198	198	100	100	100.0	41162	PASS
199	198	5	9	6.6	2703	PASS
275	198	10	60	22.6	9291	PASS
365	198	1	100	2.0	816	PASS
441	442	0.01	24	17.2	5794	PASS
442	198	50	100	82.0	33744	PASS
443	442	15	24	18.9	6384	PASS