


Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF051315\  
 Data File : BF079220.D  
 Acq On : 14 May 2015 4:52  
 Operator : TP/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF043015.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu May 14 00:58:22 2015



AutoFind: Scans 1050, 1051, 1052; Background Corrected with Scan 1047

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 36.1      | 7064    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 32.8      | 6431    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 47.9      | 9374    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 19587   | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 1321    | PASS             |
| 275         | 198          | 10           | 60           | 25.6      | 5022    | PASS             |
| 365         | 198          | 1            | 100          | 2.7       | 525     | PASS             |
| 441         | 442          | 0.01         | 24           | 15.8      | 2812    | PASS             |
| 442         | 198          | 50           | 100          | 90.6      | 17751   | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 3374    | PASS             |