

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051319\
 Data File : BF114223.D
 Acq On : 13 May 2019 12:47
 Operator : JU/SJ
 Sample : K2766-11
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P026-SS008-0206-01

Manual Integrations
 APPROVED

Sohil
 5/14/2019 8:03:00 PM

Quant Time: May 14 04:56:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	295702	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	1171103	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	565350	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	1108237	20.00	ng	0.00
76) Chrysene-d12	14.02	240	710078	20.00	ng	0.00
87) Perylene-d12	15.51	264	688344	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	916296	49.87	ng	0.00
7) Phenol-d6	6.49	99	1187981	54.66	ng	0.00
23) Nitrobenzene-d5	7.41	82	898499	43.06	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	310782	53.17	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	1343450	35.95	ng	0.00
79) Terphenyl-d14	12.95	244	1276222	37.05	ng	0.00
Target Compounds						
32) Benzoic acid	7.86	122	4712m	7.369	ng	Qvalue 100
50) Dimethylphthalate	9.60	163	322071	7.760	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051319\
Data File : BF114223.D
Acq On : 13 May 2019 12:47
Operator : JU/SJ
Sample : K2766-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P026-SS008-0206-01

Manual Integrations
APPROVED

Sohil
5/14/2019 8:03:00 PM

Quant Time: May 14 04:56:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

