

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051319\
 Data File : BF114244.D
 Acq On : 13 May 2019 21:55
 Operator : JU/SJ
 Sample : K2807-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-02-039-B

Manual Integrations
 APPROVED

Sohil
 5/14/2019 8:03:31 PM

Quant Time: May 14 08:08:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	223407	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	831548	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	375720	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	622237	20.00	ng	0.00
76) Chrysene-d12	14.02	240	476655	20.00	ng	0.00
87) Perylene-d12	15.51	264	414956	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1301442	93.75	ng	0.02
7) Phenol-d6	6.51	99	1723016	104.94	ng	0.02
23) Nitrobenzene-d5	7.42	82	1031195	69.59	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	343842	88.52	ng	0.01
45) 2-Fluorobiphenyl	9.21	172	1627520	65.52	ng	0.00
79) Terphenyl-d14	12.96	244	1326253	57.36	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	237680	8.617	ng	98
58) Fluorene	10.44	166	67661	2.620	ng	99
71) Phenanthrene	11.40	178	467651	15.448	ng	98
72) Anthracene	11.45	178	138210	4.545	ng	97
75) Fluoranthene	12.59	202	724234	22.216	ng	98
78) Pyrene	12.82	202	663020	17.291	ng	98
80) Butylbenzylphthalate	13.43	149	40140	2.487	ng	91
81) Benzo(a)anthracene	14.01	228	319844	10.375	ng	98
83) Chrysene	14.04	228	271878	8.996	ng	98
86) Indeno(1,2,3-cd)pyrene	16.99	276	89434	3.348	ng	96
88) Benzo(b)fluoranthene	15.07	252	315251m	11.858	ng	
89) Benzo(k)fluoranthene	15.10	252	102680m	4.205	ng	
90) Benzo(a)pyrene	15.44	252	212916	9.100	ng	# 95
92) Benzo(a,h,i)perylene	17.44	276	94471	4.849	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051319\
Data File : BF114244.D
Acq On : 13 May 2019 21:55
Operator : JU/SJ
Sample : K2807-03
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-GF-02-039-B

Manual Integrations
APPROVED

Sohil
5/14/2019 8:03:31 PM

Quant Time: May 14 08:08:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

