




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051324\
Data File : BF137896.D
Acq On : 13 May 2024 11:49
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 13 13:26:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration



TIC: BF137896.D\data.ms

(70) Pentachlorophenol (C)




11.375min (-0.023) 24375.43 ng

response 184173

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.10	63.70
264.00	63.70	64.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051324\
Data File : BF137896.D
Acq On : 13 May 2024 11:49
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 13 13:26:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration



TIC: BF137896.D\data.ms

(77) Benzidine

12.916min (-0.018) 0.00 ng

response 324150

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.40	14.02
183.00	11.50	11.77
0.00	0.00	0.00