

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051324\
 Data File : BF137897.D
 Acq On : 13 May 2024 12:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 13:33:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	69	-0.01
2	1,4-Dioxane	40.000	45.965	-14.9	81	-0.02
3	Pyridine	40.000	35.574	11.1	62	-0.02
4	n-Nitrosodimethylamine	40.000	38.017	5.0	67	-0.01
5 S	2-Fluorophenol	80.000	105.302	-31.6#	94	-0.01
6	Aniline	40.000	31.818	20.5	55	0.00
7 S	Phenol-d6	80.000	103.677	-29.6#	92	-0.01
8	2-Chlorophenol	40.000	38.350	4.1	68	-0.02
9	Benzaldehyde	40.000	40.333	-0.8	73	-0.02
10 C	Phenol	40.000	36.990	7.5	66	-0.01
11	bis(2-Chloroethyl)ether	40.000	37.358	6.6	66	-0.02
12	1,3-Dichlorobenzene	40.000	39.025	2.4	69	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.360	1.6	70	-0.01
14	1,2-Dichlorobenzene	40.000	38.930	2.7	69	-0.02
15	Benzyl Alcohol	40.000	38.262	4.3	65	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	33.021	17.4	58	-0.02
17	2-Methylphenol	40.000	38.722	3.2	68	-0.01
18	Hexachloroethane	40.000	39.573	1.1	69	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	36.025	9.9	64	-0.01
20	3+4-Methylphenols	40.000	38.423	3.9	67	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	68	-0.02
22	Acetophenone	40.000	38.924	2.7	67	-0.01
23 S	Nitrobenzene-d5	80.000	77.805	2.7	66	-0.01
24	Nitrobenzene	40.000	38.471	3.8	65	-0.01
25	Isophorone	40.000	37.942	5.1	66	-0.01
26 C	2-Nitrophenol	40.000	42.075	-5.2	67	-0.02
27	2,4-Dimethylphenol	40.000	38.596	3.5	66	-0.02
28	bis(2-Chloroethoxy)methane	40.000	37.913	5.2	65	-0.01
29 C	2,4-Dichlorophenol	40.000	39.912	0.2	67	-0.02
30	1,2,4-Trichlorobenzene	40.000	40.266	-0.7	69	-0.02
31	Naphthalene	40.000	39.550	1.1	68	-0.02
32	Benzoic acid	40.000	38.149	4.6	63	-0.01
33	4-Chloroaniline	40.000	34.086	14.8	58	-0.02
34 C	Hexachlorobutadiene	40.000	41.644	-4.1	71	-0.01
35	Caprolactam	40.000	38.041	4.9	64	-0.02
36 C	4-Chloro-3-methylphenol	40.000	39.653	0.9	67	-0.02
37	2-Methylnaphthalene	40.000	39.250	1.9	67	-0.02
38	1-Methylnaphthalene	40.000	38.781	3.0	67	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	67	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	40.409	-1.0	68	-0.02
41 P	Hexachlorocyclopentadiene	40.000	38.907	2.7	62	-0.01
42 S	2,4,6-Tribromophenol	80.000	109.886	-37.4#	93	-0.02
43 C	2,4,6-Trichlorophenol	40.000	38.827	2.9	64	-0.02
44	2,4,5-Trichlorophenol	40.000	39.575	1.1	65	-0.02
45 S	2-Fluorobiphenyl	80.000	79.575	0.5	69	-0.02
46	1,1'-Biphenyl	40.000	39.592	1.0	67	-0.02
47	2-Chloronaphthalene	40.000	39.304	1.7	67	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051324\
 Data File : BF137897.D
 Acq On : 13 May 2024 12:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 13:33:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.913	2.7	63	-0.01
49	Acenaphthylene	40.000	39.331	1.7	67	-0.02
50	Dimethylphthalate	40.000	38.948	2.6	66	-0.01
51	2,6-Dinitrotoluene	40.000	39.895	0.3	66	-0.02
52 C	Acenaphthene	40.000	39.361	1.6	67	-0.02
53	3-Nitroaniline	40.000	37.531	6.2	61	-0.02
54 P	2,4-Dinitrophenol	40.000	33.789	15.5	55	-0.02
55	Dibenzofuran	40.000	39.010	2.5	66	-0.02
56 P	4-Nitrophenol	40.000	35.867	10.3	58	-0.02
57	2,4-Dinitrotoluene	40.000	40.234	-0.6	66	-0.02
58	Fluorene	40.000	38.462	3.8	65	-0.02
59	2,3,4,6-Tetrachlorophenol	40.000	36.850	7.9	61	-0.02
60	Diethylphthalate	40.000	39.460	1.3	67	-0.02
61	4-Chlorophenyl-phenylether	40.000	38.955	2.6	67	-0.02
62	4-Nitroaniline	40.000	35.998	10.0	59	-0.02
63	Azobenzene	40.000	37.022	7.4	63	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	65	-0.02
65	4,6-Dinitro-2-methylphenol	40.000	38.538	3.7	60	-0.02
66 c	n-Nitrosodiphenylamine	40.000	39.778	0.6	65	-0.02
67	4-Bromophenyl-phenylether	40.000	40.471	-1.2	66	-0.02
68	Hexachlorobenzene	40.000	39.792	0.5	65	-0.02
69	Atrazine	40.000	32.706	18.2	54	-0.02
70 C	Pentachlorophenol	40.000	36.244	9.4	57	-0.02
71	Phenanthrene	40.000	38.883	2.8	65	-0.02
72	Anthracene	40.000	38.900	2.8	64	-0.02
73	Carbazole	40.000	38.150	4.6	63	-0.02
74	Di-n-butylphthalate	40.000	40.672	-1.7	67	-0.02
75 C	Fluoranthene	40.000	37.655	5.9	63	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	56	-0.02
77	Benzydine	40.000	7.428	81.4#	9	-0.02
78	Pyrene	40.000	43.352	-8.4	61	-0.02
79 S	Terphenyl-d14	80.000	88.448	-10.6	64	-0.02
80	Butylbenzylphthalate	40.000	45.772	-14.4	62	-0.02
81	Benzo(a)anthracene	40.000	39.707	0.7	56	-0.02
82	3,3'-Dichlorobenzidine	40.000	36.394	9.0	52	-0.02
83	Chrysene	40.000	39.778	0.6	57	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	46.069	-15.2	65	-0.02
85 c	Di-n-octyl phthalate	40.000	46.868	-17.2	67	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	65	-0.03
87	Indeno(1,2,3-cd)pyrene	40.000	47.169	-17.9	77	-0.04
88	Benzo(b)fluoranthene	40.000	36.367	9.1	60	-0.03
89	Benzo(k)fluoranthene	40.000	38.734	3.2	62	-0.03
90 C	Benzo(a)pyrene	40.000	39.489	1.3	65	-0.04
91	Dibenzo(a,h)anthracene	40.000	47.745	-19.4	78	-0.05
92	Benzo(g,h,i)perylene	40.000	47.753	-19.4	77	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051324\
Data File : BF137897.D
Acq On : 13 May 2024 12:18
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 13 13:33:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0