

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051415\
 Data File : BF079240.D
 Acq On : 14 May 2015 18:51
 Operator : TP/IZ
 Sample : G2215-12
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-39D

Quant Time: May 15 03:28:28 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 02:17:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.22	152	50792	20.00	ng	0.00
21) Naphthalene-d8	8.79	136	212084	20.00	ng	-0.01
38) Acenaphthene-d10	10.97	164	106013	20.00	ng	0.00
63) Phenanthrene-d10	12.81	188	203981	20.00	ng	0.00
75) Chrysene-d12	16.14	240	225121	20.00	ng	-0.10
86) Perylene-d12	18.02	264	251917	20.00	ng	-0.27

System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	237561m	80.51	ng	0.00
7) Phenol-d6	6.79	99	321677	82.48	ng	0.00
23) Nitrobenzene-d5	7.91	82	186911	50.65	ng	0.00
41) 2,4,6-Tribromophenol	11.95	330	94613	82.50	ng	0.00
44) 2-Fluorobiphenyl	10.14	172	334931	44.02	ng	-0.01
78) Terphenyl-d14	14.81	244	378787	40.28	ng	-0.02

Target Compounds				Qvalue		
31) Naphthalene	8.82	128	50841	5.03	ng	98
37) 2-Methylnaphthalene	9.68	142	44139	6.30	ng	98
51) Acenaphthene	11.01	154	43596	6.77	ng	100
54) Dibenzofuran	11.22	168	28935	3.11	ng	94
57) Fluorene	11.65	166	62150	8.14	ng	98
70) Phenanthrene	12.85	178	604180	52.94	ng	100
71) Anthracene	12.90	178	161591	14.43	ng	99
72) Carbazole	13.11	167	35659	3.34	ng	99
74) Fluoranthene	14.32	202	737296	62.00	ng	99
77) Pyrene	14.61	202	774470	59.70	ng	100
80) Benzo(a)anthracene	16.13	228	360847	29.27	ng	95
82) Chrysene	16.17	228	316893	26.73	ng	97
85) Indeno(1,2,3-cd)pyrene	19.49	276	209258m	13.85	ng	
87) Benzo(b)fluoranthene	17.53	252	423200m	30.69	ng	
88) Benzo(k)fluoranthene	17.57	252	149930m	11.23	ng	
89) Benzo(a)pyrene	17.95	252	334610	26.35	ng	98
90) Dibenzo(a,h)anthracene	19.51	278	51818m	3.84	ng	
91) Benzo(g,h,i)perylene	19.91	276	205237	15.18	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051415\
Data File : BF079240.D
Acq On : 14 May 2015 18:51
Operator : TP/IZ
Sample : G2215-12
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-39D

Quant Time: May 15 03:28:28 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 02:17:41 2015
Response via : Initial Calibration

