

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051415\
 Data File : BF079251.D
 Acq On : 15 May 2015 00:50
 Operator : TP/IZ
 Sample : G2246-02 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUP-1

Quant Time: May 15 04:12:16 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 02:17:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.22	152	44151	20.00	ng	0.00
21) Naphthalene-d8	8.80	136	177473	20.00	ng	0.00
38) Acenaphthene-d10	10.97	164	88319	20.00	ng	0.00
63) Phenanthrene-d10	12.81	188	178378	20.00	ng	0.00
75) Chrysene-d12	16.10	240	195554	20.00	ng	-0.14
86) Perylene-d12	17.82	264	202796	20.00	ng	-0.48
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	160134m	62.43	ng	0.00
7) Phenol-d6	6.79	99	191464	56.47	ng	0.00
23) Nitrobenzene-d5	7.91	82	109347	35.41	ng	0.00
41) 2,4,6-Tribromophenol	11.95	330	55257	57.83	ng	0.00
44) 2-Fluorobiphenyl	10.14	172	232726	36.71	ng	-0.01
78) Terphenyl-d14	14.81	244	278101m	34.05	ng	-0.02
Target Compounds						Qvalue
70) Phenanthrene	12.83	178	48853	4.90	ng	98
74) Fluoranthene	14.32	202	105079	10.11	ng	94
77) Pyrene	14.59	202	105464m	9.36	ng	
80) Benzo(a)anthracene	16.08	228	63339m	5.91	ng	
82) Chrysene	16.13	228	56358	5.47	ng	98
85) Indeno(1,2,3-cd)pyrene	19.23	276	47150m	3.59	ng	
87) Benzo(b)fluoranthene	17.38	252	87373m	7.87	ng	
88) Benzo(k)fluoranthene	17.41	252	32066m	2.98	ng	
89) Benzo(a)pyrene	17.75	252	67925	6.65	ng	# 93
91) Benzo(g,h,i)perylene	19.66	276	46880	4.31	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051415\
Data File : BF079251.D
Acq On : 15 May 2015 00:50
Operator : TP/IZ
Sample : G2246-02 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DUP-1

Quant Time: May 15 04:12:16 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 02:17:41 2015
Response via : Initial Calibration

