


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051416\
 Data File : BF087050.D
 Acq On : 15 May 2016 5:51
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050916.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 13 14:55:23 2016



AutoFind: Scans 857, 858, 859; Background Corrected with Scan 854

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	71.7	29226	PASS
68	69	0.00	2	1.8	351	PASS
69	198	0.00	100	49.0	19974	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.1	22471	PASS
197	198	0.00	2	0.9	355	PASS
198	198	100	100	100.0	40786	PASS
199	198	5	9	6.8	2778	PASS
275	198	10	60	24.8	10129	PASS
365	198	1	100	3.9	1596	PASS
441	442	0.01	24	15.7	5862	PASS
442	198	50	100	91.3	37250	PASS
443	442	15	24	18.1	6730	PASS