

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051420\  
 Data File : BF120302.D  
 Acq On : 14 May 2020 20:33  
 Operator : MA/SJ  
 Sample : L2635-01  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ROLL-OFF-P72-11933

Quant Time: May 15 04:06:17 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu May 14 12:47:38 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.62  | 152  | 280835   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.90  | 136  | 1167022  | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.66  | 164  | 575738   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.14 | 188  | 961172   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 13.78 | 240  | 560757   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 15.17 | 264  | 571240   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.23  | 112  | 1802486  | 124.88 | ng    | 0.00     |
| 7) Phenol-d6                | 6.27  | 99   | 2121404  | 116.22 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.19  | 82   | 1353860  | 79.43  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.45 | 330  | 528156   | 112.51 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 8.99  | 172  | 2552264  | 91.08  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.73 | 244  | 2370382  | 93.39  | ng    | 0.00     |
| Target Compounds            |       |      |          |        |       |          |
| 10) Phenol                  | 6.29  | 94   | 47139    | 2.212  | ng    | 89       |
| 50) Dimethylphthalate       | 9.39  | 163  | 744893   | 20.467 | ng    | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051420\  
Data File : BF120302.D  
Acq On : 14 May 2020 20:33  
Operator : MA/SJ  
Sample : L2635-01  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
ROLL-OFF-P72-11933

Quant Time: May 15 04:06:17 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 14 12:47:38 2020  
Response via : Initial Calibration

